



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: M1785

MATRIX: Water

METHOD: 8260-Low

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The % RSD is greater than 15% in the Initial Calibration method (82X032421W.M) for Chloromethane and 1,2-Dibromo-3-Chloropropane are passing on Linear regression. The Continuous Calibration met the requirements .		✓	
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

	NA	NO	YES
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The MS {M1785-11MS} with File ID: VX021589.D recoveries met the requirements for all compounds except for 1,2,4-Trimethylbenzene [58%] and N-propylbenzene[76%] .			
The MSD {M1785-12MSD} with File ID: VX021590.D recoveries met the acceptable requirements except for 1,2,4-Trimethylbenzene[58%] and N-propylbenzene[74%].			
The RPD met criteria. due to high concentration of original sample MS/MSD recovery outside the QC limit..			
9. Internal Standard Area/Retention Time Shift Meet Criteria			✓
Comments:			
10. Analysis Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

This data Package has been revised due to TICS corrected for sample # 14.

Samples MW-1, MW-2, MW-3R, MW-4, MW-5R and FD-032321 the samples are analyzed at straight dilution due to high contamination of analyte 124-trimethyl benzene.

Sample MW-3R was diluted due to high concentration.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date