



## Cover Page

**Order ID :** M1785

**Project ID :** Ebenzer Plaza-1 - 20918.10

**Client :** Chazen Companies

### Lab Sample Number

M1785-10  
M1785-11  
M1785-12  
M1785-13  
M1785-14  
M1785-15  
M1785-16  
M1785-17  
M1785-18

### Client Sample Number

MW-1  
MW-1MS  
MW-1MSD  
MW-2  
MW-3R  
MW-4  
MW-5R  
FD-032321  
TRIP-BLANK

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : \_\_\_\_\_

Date: 4/5/2021



NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION  
FORM S-1

SAMPLE IDENTIFICATION AND ANALYTICAL REQUIREMENT SUMMARY

NYSDEC Sample ID/Code	Laboratory Sampl ID/Code	VOA GC/MS (Method #)	BNA GC/MS (Method #)	VOA GC (Method #)	Pest PCBs (Method #)	Metals (Method #)	Other (Method #)
MW-1	M1785-10	8260-Low					
MW-2	M1785-13	8260-Low					
MW-3R	M1785-14	8260-Low					
MW-4	M1785-15	8260-Low					
MW-5R	M1785-16	8260-Low					
FD-032321	M1785-17	8260-Low					
TRIP-BLANK	M1785-18	8260-Low					



NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION

FORM S-IIb

SAMPLE PREPARATION AND ANALYSIS SUMMARY VOLATILE (VOA) ANALYSES

Laboratory Sample ID	Matrix	Date Collected	Date Rec'd at Lab	Date Extracted	Date Analyzed
M1785-10	Water	03/23/21	03/23/21		03/29/21
M1785-13	Water	03/23/21	03/23/21		03/29/21
M1785-14	Water	03/23/21	03/23/21		03/29/21
M1785-15	Water	03/23/21	03/23/21		03/29/21
M1785-16	Water	03/23/21	03/23/21		03/29/21
M1785-17	Water	03/23/21	03/23/21		03/29/21
M1785-18	Water	03/23/21	03/23/21		03/29/21

\* Details For Test : VOCMS Group2



NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION

FORM S-III

SAMPLE PREPARATION AND ANALYSIS SUMMARY MISCELLANEOUS ORGANIC ANALYSES

Laboratory Sample ID	Matrix	Analytical Protocol	Extraction Method	Auxiliary Cleanup	Dil/Conc Factor
M1785-10	Water	8260-Low	5030		
M1785-11	Water	8260-Low	5030		
M1785-12	Water	8260-Low	5030		
M1785-13	Water	8260-Low	5030		
M1785-14	Water	8260-Low	5030		
M1785-15	Water	8260-Low	5030		
M1785-16	Water	8260-Low	5030		
M1785-17	Water	8260-Low	5030		
M1785-18	Water	8260-Low	5030		



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Chazen Companies**

**Project Name: Ebenzer Plaza-1 - 20918.10**

**Project # N/A**

**Chemtech Project # M1785**

**Test Name: VOCMS Group2**

### **A. Number of Samples and Date of Receipt:**

9 Water samples were received on 03/23/2021.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: VOCMS Group2. This data package contains results for VOCMS Group2.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOCMS Group2 was based on method 8260-Low.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {M1785-11MS} with File ID: VX021589.D recoveries met the requirements for all compounds except for 1,2,4-Trimethylbenzene [58%] and N-propylbenzene[76%] .

The MSD {M1785-12MSD} with File ID: VX021590.D recoveries met the acceptable requirements except for 1,2,4-Trimethylbenzene[58%] and N-propylbenzene[74%].

The RPD met criteria. due to high concentration of original sample MS/MSD recovery outside the QC limit.

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 15% in the Initial Calibration method (82X032421W.M) for Chloromethane and 1,2-Dibromo-3-Chloropropane are passing on Linear regression.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Samples MW-1, MW-2, MW-3R, MW-4, MW-5R and FD-032321 the samples are analyzed at straight dilution due to high contamination of analyte 124-trimethyl benzene. Sample MW-3R was diluted due to high concentration.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

**E. Additional Comments:**

This data Package has been revised due to TICS corrected for sample # 14.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_

## DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

**APPENDIX A****QA REVIEW GENERAL DOCUMENTATION**

Project #: M1785

Completed

---

For thorough review, the report must have the following:

**GENERAL:**

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

---

☒

Check chain-of-custody for proper relinquish/return of samples

---

☒

Is the chain of custody signed and complete

---

☒

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

---

☒

Collect information for each project id from server. Were all requirements followed

---

☒**COVER PAGE:**

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

---

☒

Do lab numbers and client Ids on cover page agree with the Chain of Custody

---

☒**CHAIN OF CUSTODY:**

Do requested analyses on Chain of Custody agree with form I results

---

☒

Do requested analyses on Chain of Custody agree with the log-in page

---

☒

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

---

☒

Were the samples received within hold time

---

☒

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

---

☒**ANALYTICAL:**

Was method requirement followed?

---

☒

Was client requirement followed?

---

☒

Does the case narrative summarize all QC failure?

---

☒

All runlogs and manual integration are reviewed for requirements

---

☒

All manual calculations and /or hand notations verified

---

☒

---

1st Level QA Review Signature: SOHIL JODHANI

Date: 04/05/2021

2nd Level QA Review Signature: \_\_\_\_\_

Date: \_\_\_\_\_