


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032921\
 Data File : VX021563.D
 Acq On : 29 Mar 2021 09:02
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Title : SW846 8260
 Last Update : Wed Mar 24 12:25:26 2021



AutoFind: Scans 1704, 1705, 1706; Background Corrected with Scan 1698

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.8	4200	PASS
75	95	30	60	55.8	10266	PASS
95	95	100	100	100.0	18410	PASS
96	95	5	9	7.1	1308	PASS
173	174	0.00	2	0.6	75	PASS
174	95	50	100	71.4	13141	PASS
175	174	5	9	8.3	1090	PASS
176	174	95	101	96.8	12718	PASS
177	176	5	9	6.7	852	PASS