



Surrogate Summary

SDG No.: M1785

Client: Chazen Companies

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
M1785-10	MW-1	1,2-Dichloroethane-d4	50	50.6	101	78	117
		Dibromofluoromethane	50	50.1	100	75	124
		Toluene-d8	50	50.1	100	92	112
		4-Bromofluorobenzene	50	52.4	105	83	123
M1785-11MS	MW-1MS	1,2-Dichloroethane-d4	50	50.9	102	78	117
		Dibromofluoromethane	50	52.6	105	75	124
		Toluene-d8	50	53.1	106	92	112
		4-Bromofluorobenzene	50	54.4	109	83	123
M1785-12MSD	MW-1MSD	1,2-Dichloroethane-d4	50	52.8	105	78	117
		Dibromofluoromethane	50	50.9	102	75	124
		Toluene-d8	50	51.5	103	92	112
		4-Bromofluorobenzene	50	52.3	105	83	123
M1785-13	MW-2	1,2-Dichloroethane-d4	50	52.5	105	78	117
		Dibromofluoromethane	50	50.6	101	75	124
		Toluene-d8	50	51.1	102	92	112
		4-Bromofluorobenzene	50	59.1	118	83	123
M1785-14	MW-3R	1,2-Dichloroethane-d4	50	50.9	102	78	117
		Dibromofluoromethane	50	50.0	100	75	124
		Toluene-d8	50	50.1	100	92	112
		4-Bromofluorobenzene	50	52.8	106	83	123
M1785-14DL	MW-3RDL	1,2-Dichloroethane-d4	50	51.5	103	78	117
		Dibromofluoromethane	50	50.6	101	75	124
		Toluene-d8	50	50.4	101	92	112
		4-Bromofluorobenzene	50	56.9	114	83	123
M1785-15	MW-4	1,2-Dichloroethane-d4	50	51.0	102	78	117
		Dibromofluoromethane	50	50.2	100	75	124
		Toluene-d8	50	50.2	100	92	112
		4-Bromofluorobenzene	50	56.1	112	83	123
M1785-16	MW-5R	1,2-Dichloroethane-d4	50	52.4	105	78	117
		Dibromofluoromethane	50	50.4	101	75	124
		Toluene-d8	50	51.2	102	92	112
		4-Bromofluorobenzene	50	57.5	115	83	123
M1785-17	FD-032321	1,2-Dichloroethane-d4	50	51.8	104	78	117
		Dibromofluoromethane	50	50.8	102	75	124
		Toluene-d8	50	48.9	98	92	112
		4-Bromofluorobenzene	50	51.5	103	83	123
M1785-18	TRIP-BLANK	1,2-Dichloroethane-d4	50	51.7	103	78	117
		Dibromofluoromethane	50	51.1	102	75	124
		Toluene-d8	50	50.4	101	92	112
		4-Bromofluorobenzene	50	53.1	106	83	123
VX0329WBL01	VX0329WBL01	1,2-Dichloroethane-d4	50	50.4	101	78	117
		Dibromofluoromethane	50	50.0	100	75	124
		Toluene-d8	50	50.1	100	92	112
		4-Bromofluorobenzene	50	53.7	107	83	123
VX0329WBS01	VX0329WBS01	1,2-Dichloroethane-d4	50	49.8	100	78	117
		Dibromofluoromethane	50	50.6	101	75	124
		Toluene-d8	50	50.9	102	92	112
		4-Bromofluorobenzene	50	49.9	100	83	123