



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHAZ02
Lab Code: CHEM Case No.: M1785 SAS No.: M1785 SDG NO.: M1785
Lab File ID: VX021564.D Date Analyzed: 03/29/2021
Instrument ID: MSVOA_X Time Analyzed: 09:35
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	70614	12.059				
UPPER LIMIT	141228	12.559				
LOWER LIMIT	35307	11.559				
EPA SAMPLE NO.						
MW-1	48214	12.06				
MW-1MS	60876	12.06				
MW-1MSD	62506	12.06				
MW-2	57707	12.06				
MW-3R	50223	12.07				
MW-3RDL	59501	12.06				
MW-4	58057	12.06				
MW-5R	57549	12.06				
FD-032321	50719	12.07				
TRIP-BLANK	47890	12.06				
VX0329WBL01	57784	12.06				
VX0329WBS01	65643	12.06				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.