



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC042222AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC059373.D

Aliphatic C9-C12	7416761.000	60.000	2.862	6.099	123612.683	127377.051	2.955
Aliphatic C12-C16	5347385.000	40.000	6.100	9.471	133684.625	137853.577	3.024
Aliphatic C16-C21	8570500.000	60.000	9.472	12.813	142841.667	149386.165	4.381
Aliphatic C21-C28	11228095.000	80.000	12.814	16.454	140351.188	154246.896	9.009
Aliphatic C28-C40	11827992.000	120.000	16.455	21.124	98566.600	117535.712	16.139
Aliphatic EPH	44390733.000	360.000	2.862	21.124	123307.592	134899.925	8.593