



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC042222AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC059395.D

Aliphatic C9-C12	7720386.000	60.000	2.862	6.099	128673.100	127377.051	-1.017
Aliphatic C12-C16	5634325.000	40.000	6.100	9.471	140858.125	137853.577	-2.180
Aliphatic C16-C21	8826621.000	60.000	9.472	12.813	147110.350	149386.165	1.523
Aliphatic C21-C28	11056145.000	80.000	12.814	16.454	138201.813	154246.896	10.402
Aliphatic C28-C40	11972270.000	120.000	16.455	21.124	99768.917	117535.712	15.116
Aliphatic EPH	45209747.000	360.000	2.862	21.124	125582.631	134899.925	6.907