



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH QC BATCHID: LB120604

	NA	NO	YES
1. Blank Contamination If YES, list compounds and concentrations in each blank:	_____	<u>✓</u>	_____
2.Matrix Spike Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside of the acceptance range: MS and MSD didn't meet criteria due to matrix interferences	_____	<u>✓</u>	_____
3. Sample Duplicate Analysis Meet QC Criteria: If not met, list those compounds and their % differences which fall outside of the acceptance range:	_____	_____	<u>✓</u>
4.Analysis Holding Time Met : If not met, list number of days exceeded for each sample:	_____	_____	<u>✓</u>

ADDITIONAL COMMENTS:

N3259-01 was analyzed with 10X dilution due to brown matrix and client history.

Analyst Guerda

Date 06/13/2022

QA REVIEW

Date

QA Control # A3040188