



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: Inframark

SDG No.: N3259

Project: Fresh Kills Landfill Leachate Treatment Plant

RunNo.: LB120595

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Phosphorus, Total	mg/L	< 0.0250	0.0250	U	0.0096	0.05	06/09/2022
Sample ID: CCB1 Phosphorus, Total	mg/L	< 0.0250	0.0250	U	0.0096	0.05	06/09/2022
Sample ID: CCB2 Phosphorus, Total	mg/L	< 0.0250	0.0250	U	0.0096	0.05	06/09/2022
Sample ID: CCB3 Phosphorus, Total	mg/L	< 0.0250	0.0250	U	0.0096	0.05	06/09/2022

Initial and Continuing Calibration Blank Summary

Client: Inframark

SDG No.: N3259

Project: Fresh Kills Landfill Leachate Treatment Plant

RunNo.: LB120604

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.14	2	06/06/2022
Chloride	mg/L	< 0.3000	0.3000	U	0.097	0.6	06/06/2022
Fluoride	mg/L	< 0.2000	0.2000	U	0.092	0.4	06/06/2022
Nitrite	mg/L	< 0.3000	0.3000	U	0.032	0.6	06/06/2022
Nitrate	mg/L	< 0.2500	0.2500	U	0.050	0.5	06/06/2022
Sulfate	mg/L	< 1.5000	1.5000	U	0.18	3	06/06/2022
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.30	1	06/06/2022
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.14	2	06/09/2022
Chloride	mg/L	< 0.3000	0.3000	U	0.097	0.6	06/09/2022
Fluoride	mg/L	< 0.2000	0.2000	U	0.092	0.4	06/09/2022
Nitrite	mg/L	< 0.3000	0.3000	U	0.032	0.6	06/09/2022
Nitrate	mg/L	< 0.2500	0.2500	U	0.050	0.5	06/09/2022
Sulfate	mg/L	< 1.5000	1.5000	U	0.18	3	06/09/2022
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.30	1	06/09/2022
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.14	2	06/09/2022
Chloride	mg/L	< 0.3000	0.3000	U	0.097	0.6	06/09/2022
Fluoride	mg/L	< 0.2000	0.2000	U	0.092	0.4	06/09/2022
Nitrite	mg/L	< 0.3000	0.3000	U	0.032	0.6	06/09/2022
Nitrate	mg/L	< 0.2500	0.2500	U	0.050	0.5	06/09/2022
Sulfate	mg/L	< 1.5000	1.5000	U	0.18	3	06/09/2022
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.30	1	06/09/2022
Sample ID: CCB3							
Bromide	mg/L	< 1.0000	1.0000	U	0.14	2	06/09/2022
Chloride	mg/L	< 0.3000	0.3000	U	0.097	0.6	06/09/2022
Fluoride	mg/L	< 0.2000	0.2000	U	0.092	0.4	06/09/2022
Nitrite	mg/L	< 0.3000	0.3000	U	0.032	0.6	06/09/2022
Nitrate	mg/L	< 0.2500	0.2500	U	0.050	0.5	06/09/2022
Sulfate	mg/L	< 1.5000	1.5000	U	0.18	3	06/09/2022
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.30	1	06/09/2022

Initial and Continuing Calibration Blank Summary**Client:** Inframark**SDG No.:** N3259**Project:** Fresh Kills Landfill Leachate Treatment Plant**RunNo.:** LB120613

Analyte		Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID:	ICB							
COD		mg/L	< 5.0000	5.0000	U	2.40	10	05/19/2022
Sample ID:	CCB1							
COD		mg/L	< 5.0000	5.0000	U	2.40	10	06/10/2022
Sample ID:	CCB2							
COD		mg/L	< 5.0000	5.0000	U	2.40	10	06/10/2022
Sample ID:	CCB3							
COD		mg/L	< 5.0000	5.0000	U	2.40	10	06/10/2022



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Initial and Continuing Calibration Blank Summary

Client:	Inframark	SDG No.:	N3259
Project:	Fresh Kills Landfill Leachate Treatment Plant	RunNo.:	LB120613

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------