



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Duplicate Sample Summary

Client:	Inframark	SDG No.:	N3259
Project:	Fresh Kills Landfill Leachate Treatment Plant	Sample ID:	LB120567BS
Client ID:	LB120567BSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	16.9		16.7		1	1.19		06/08/2022



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Duplicate Sample Summary

Client:	Inframark	SDG No.:	N3259
Project:	Fresh Kills Landfill Leachate Treatment Plant	Sample ID:	N3252-01
Client ID:	POOL-WATERMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Nitrate	mg/L	+/-20	2.80		3.50		1	22	*	06/09/2022
Bromide	mg/L	+/-20	10.2		12.9		1	23	*	06/09/2022
Orthophosphate as P	mg/L	+/-20	6.60		8.30		1	23	*	06/09/2022
Nitrite	mg/L	+/-20	3.00		3.80		1	24	*	06/09/2022
Sulfate	mg/L	+/-20	15.3		19.4		1	24	*	06/09/2022
Chloride	mg/L	+/-20	3.20		4.10		1	25	*	06/09/2022
Fluoride	mg/L	+/-20	2.20		2.90		1	27	*	06/09/2022



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Duplicate Sample Summary

Client:	Inframark	SDG No.:	N3259
Project:	Fresh Kills Landfill Leachate Treatment Plant	Sample ID:	N3261-01DL
Client ID:	BI-WEEKLY-DECANTDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/L	+/-20	1.10	D	1.09	D	5	0.91		06/09/2022



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Duplicate Sample Summary

Client:	Inframark	SDG No.:	N3259
Project:	Fresh Kills Landfill Leachate Treatment Plant	Sample ID:	N3261-01DL
Client ID:	BI-WEEKLY-DECANTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/L	+/-20	1.71	D	1.71	D	5	0		06/09/2022



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Duplicate Sample Summary

Client:	Inframark	SDG No.:	N3259
Project:	Fresh Kills Landfill Leachate Treatment Plant	Sample ID:	N3263-01
Client ID:	DSW002DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	21.5		22.5		1	4.55		06/10/2022



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Duplicate Sample Summary

Client:	Inframark	SDG No.:	N3259
Project:	Fresh Kills Landfill Leachate Treatment Plant	Sample ID:	N3263-01
Client ID:	DSN002MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	70.9		70.9		1	0		06/10/2022