

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH QC BATCHID: LB120624

	NA	NO	YES
1.Calibration Summary met criteria.	_____	_____	_____✓_____
2.ICP Interference Check Sample Results Summary Submitted.	_____	_____	_____✓_____
3. Serial Dilution Summary (if applicable) Submitted.	_____	_____	_____✓_____
4.Laboratory Control Sample Summary (if applicable) Submitted.:	_____	_____	_____✓_____
5.Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓_____	_____
6.Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
7.Sample Duplicate Analysis Met QC Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
8.Digestion Holding Time Met If not met, list number of days exceeded for each sample:	_____	_____	_____✓_____
9.Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____

ADDITIONAL COMMENTS:

N3270-06 MS/MSD fail for (Ca,Cr,Mn,K,Na,V) due to both chemical interference during digestion process and unknow sample matrix interference. Post spike recovery still fail for 2 elements.

Analyst bin

Date 6/14/2022 12:00:00 AM

QA REVIEW

Date _____