



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH QC BATCHID: LB120705

	NA	NO	YES
1.Calibration Summary met criteria.	_____	_____	<u>✓</u>
2.ICP Interference Check Sample Results Summary Submitted.	_____	_____	<u>✓</u>
3. Serial Dilution Summary (if applicable) Submitted.	_____	_____	<u>✓</u>
4.Laboratory Control Sample Summary (if applicable) Submitted.:	_____	_____	<u>✓</u>
5.Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	<u>✓</u>	_____
6.Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	<u>✓</u>
7.Sample Duplicate Analysis Met QC Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	<u>✓</u>
8.Digestion Holding Time Met If not met, list number of days exceeded for each sample:	_____	_____	<u>✓</u>
9.Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	<u>✓</u>

ADDITIONAL COMMENTS:

MS/MSD fail for N3340-03 (Ca) parameters due to chemical interference during digestion process. Post spike recovery passed.

Analyst bin

Date 6/16/2022 12:00:00 AM

QA REVIEW

Date