



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Inframark **SDG No.:** N3259
Contract: INFR01 **Lab Code:** CHEM **Case No.:** N3259 **SAS No.:** N3259
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	110	100	110	40 - 160	P	06/10/2022	16:05	LB120624
	Arsenic	20.6	20.0	103	40 - 160	P	06/10/2022	16:05	LB120624
	Copper	23.1	20.0	115	40 - 160	P	06/10/2022	16:05	LB120624
	Iron	111	100	111	40 - 160	P	06/10/2022	16:05	LB120624
	Zinc	36.5	40.0	91	40 - 160	P	06/10/2022	16:05	LB120624
CRI01	Aluminum	105	100	105	40 - 160	P	06/14/2022	15:09	LB120678
	Arsenic	20.7	20.0	104	40 - 160	P	06/14/2022	15:09	LB120678
	Copper	21.8	20.0	109	40 - 160	P	06/14/2022	15:09	LB120678
	Iron	102	100	102	40 - 160	P	06/14/2022	15:09	LB120678
	Zinc	36.0	40.0	90	40 - 160	P	06/14/2022	15:09	LB120678
CRI01	Aluminum	100	100	100	40 - 160	P	06/15/2022	14:47	LB120705
	Arsenic	19.3	20.0	97	40 - 160	P	06/15/2022	14:47	LB120705
	Copper	20.2	20.0	101	40 - 160	P	06/15/2022	14:47	LB120705
	Iron	93.6	100	94	40 - 160	P	06/15/2022	14:47	LB120705
	Zinc	37.0	40.0	93	40 - 160	P	06/15/2022	14:47	LB120705