



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Inframark **SDG No.:** N3259
Contract: INFR01 **Lab Code:** CHEM **Case No.:** N3259 **SAS No.:** N3259

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	06/10/2022	16:01	LB120624
	Arsenic	20.0	+/-20.0	U	20.0	P	06/10/2022	16:01	LB120624
	Copper	20.0	+/-20.0	U	20.0	P	06/10/2022	16:01	LB120624
	Iron	100	+/-100	U	100	P	06/10/2022	16:01	LB120624
	Zinc	40.0	+/-40.0	U	40.0	P	06/10/2022	16:01	LB120624



Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Inframark				SDG No.:	N3259			
Contract:	INFR01	Lab Code:	CHEM		Case No.:	N3259	SAS No.:	N3259	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	06/10/2022	16:33	LB120624
	Arsenic	20.0	+/-20.0	U	20.0	P	06/10/2022	16:33	LB120624
	Copper	20.0	+/-20.0	U	20.0	P	06/10/2022	16:33	LB120624
	Iron	100	+/-100	U	100	P	06/10/2022	16:33	LB120624
	Zinc	40.0	+/-40.0	U	40.0	P	06/10/2022	16:33	LB120624
CCB02	Aluminum	100	+/-100	U	100	P	06/10/2022	17:19	LB120624
	Arsenic	20.0	+/-20.0	U	20.0	P	06/10/2022	17:19	LB120624
	Copper	20.0	+/-20.0	U	20.0	P	06/10/2022	17:19	LB120624
	Iron	100	+/-100	U	100	P	06/10/2022	17:19	LB120624
	Zinc	40.0	+/-40.0	U	40.0	P	06/10/2022	17:19	LB120624
CCB03	Aluminum	100	+/-100	U	100	P	06/10/2022	18:09	LB120624
	Arsenic	20.0	+/-20.0	U	20.0	P	06/10/2022	18:09	LB120624
	Copper	20.0	+/-20.0	U	20.0	P	06/10/2022	18:09	LB120624
	Iron	100	+/-100	U	100	P	06/10/2022	18:09	LB120624
	Zinc	40.0	+/-40.0	U	40.0	P	06/10/2022	18:09	LB120624
CCB04	Aluminum	100	+/-100	U	100	P	06/10/2022	19:00	LB120624
	Arsenic	20.0	+/-20.0	U	20.0	P	06/10/2022	19:00	LB120624
	Copper	20.0	+/-20.0	U	20.0	P	06/10/2022	19:00	LB120624
	Iron	100	+/-100	U	100	P	06/10/2022	19:00	LB120624
	Zinc	40.0	+/-40.0	U	40.0	P	06/10/2022	19:00	LB120624
CCB05	Aluminum	100	+/-100	U	100	P	06/10/2022	19:50	LB120624
	Arsenic	20.0	+/-20.0	U	20.0	P	06/10/2022	19:50	LB120624
	Copper	20.0	+/-20.0	U	20.0	P	06/10/2022	19:50	LB120624
	Iron	100	+/-100	U	100	P	06/10/2022	19:50	LB120624
	Zinc	40.0	+/-40.0	U	40.0	P	06/10/2022	19:50	LB120624
CCB06	Aluminum	100	+/-100	U	100	P	06/10/2022	20:28	LB120624
	Arsenic	20.0	+/-20.0	U	20.0	P	06/10/2022	20:28	LB120624
	Copper	20.0	+/-20.0	U	20.0	P	06/10/2022	20:28	LB120624
	Iron	100	+/-100	U	100	P	06/10/2022	20:28	LB120624
	Zinc	40.0	+/-40.0	U	40.0	P	06/10/2022	20:28	LB120624

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Inframark

SDG No.: N3259

Contract: INFR01

Lab Code: CHEM

Case No.: N3259

SAS No.: N3259

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	06/14/2022	15:05	LB120678
	Arsenic	20.0	+/-20.0	U	20.0	P	06/14/2022	15:05	LB120678
	Copper	20.0	+/-20.0	U	20.0	P	06/14/2022	15:05	LB120678
	Iron	100	+/-100	U	100	P	06/14/2022	15:05	LB120678
	Zinc	40.0	+/-40.0	U	40.0	P	06/14/2022	15:05	LB120678



Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Inframark SDG No.: N3259
Contract: INFR01 Lab Code: CHEM Case No.: N3259 SAS No.: N3259

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	06/14/2022	15:36	LB120678
	Arsenic	20.0	+/-20.0	U	20.0	P	06/14/2022	15:36	LB120678
	Copper	20.0	+/-20.0	U	20.0	P	06/14/2022	15:36	LB120678
	Iron	100	+/-100	U	100	P	06/14/2022	15:36	LB120678
	Zinc	40.0	+/-40.0	U	40.0	P	06/14/2022	15:36	LB120678
CCB02	Aluminum	100	+/-100	U	100	P	06/14/2022	16:30	LB120678
	Arsenic	20.0	+/-20.0	U	20.0	P	06/14/2022	16:30	LB120678
	Copper	20.0	+/-20.0	U	20.0	P	06/14/2022	16:30	LB120678
	Iron	100	+/-100	U	100	P	06/14/2022	16:30	LB120678
	Zinc	40.0	+/-40.0	U	40.0	P	06/14/2022	16:30	LB120678
CCB03	Aluminum	100	+/-100	U	100	P	06/14/2022	17:19	LB120678
	Arsenic	20.0	+/-20.0	U	20.0	P	06/14/2022	17:19	LB120678
	Copper	20.0	+/-20.0	U	20.0	P	06/14/2022	17:19	LB120678
	Iron	100	+/-100	U	100	P	06/14/2022	17:19	LB120678
	Zinc	40.0	+/-40.0	U	40.0	P	06/14/2022	17:19	LB120678
CCB04	Aluminum	100	+/-100	U	100	P	06/14/2022	18:21	LB120678
	Arsenic	20.0	+/-20.0	U	20.0	P	06/14/2022	18:21	LB120678
	Copper	20.0	+/-20.0	U	20.0	P	06/14/2022	18:21	LB120678
	Iron	100	+/-100	U	100	P	06/14/2022	18:21	LB120678
	Zinc	40.0	+/-40.0	U	40.0	P	06/14/2022	18:21	LB120678
CCB05	Aluminum	100	+/-100	U	100	P	06/14/2022	19:11	LB120678
	Arsenic	20.0	+/-20.0	U	20.0	P	06/14/2022	19:11	LB120678
	Copper	20.0	+/-20.0	U	20.0	P	06/14/2022	19:11	LB120678
	Iron	100	+/-100	U	100	P	06/14/2022	19:11	LB120678
	Zinc	40.0	+/-40.0	U	40.0	P	06/14/2022	19:11	LB120678
CCB06	Aluminum	100	+/-100	U	100	P	06/14/2022	19:35	LB120678
	Arsenic	20.0	+/-20.0	U	20.0	P	06/14/2022	19:35	LB120678
	Copper	20.0	+/-20.0	U	20.0	P	06/14/2022	19:35	LB120678
	Iron	100	+/-100	U	100	P	06/14/2022	19:35	LB120678
	Zinc	40.0	+/-40.0	U	40.0	P	06/14/2022	19:35	LB120678



- 3a -

Client:	Inframark	SDG No.:	N3259				
Contract:	INFR01	Lab Code:	CHEM	Case No.:	N3259	SAS No.:	N3259

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	06/15/2022	14:43	LB120705
	Arsenic	20.0	+/-20.0	U	20.0	P	06/15/2022	14:43	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	14:43	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	14:43	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	14:43	LB120705



Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Inframark				SDG No.:	N3259			
Contract:	INFR01	Lab Code:	CHEM			Case No.:	N3259	SAS No.:	N3259
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	06/15/2022	15:22	LB120705
	Arsenic	20.0	+/-20.0	U	20.0	P	06/15/2022	15:22	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	15:22	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	15:22	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	15:22	LB120705
CCB02	Aluminum	100	+/-100	U	100	P	06/15/2022	16:08	LB120705
	Arsenic	20.0	+/-20.0	U	20.0	P	06/15/2022	16:08	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	16:08	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	16:08	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	16:08	LB120705
CCB03	Aluminum	100	+/-100	U	100	P	06/15/2022	16:54	LB120705
	Arsenic	20.0	+/-20.0	U	20.0	P	06/15/2022	16:54	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	16:54	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	16:54	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	16:54	LB120705
CCB04	Aluminum	100	+/-100	U	100	P	06/15/2022	17:40	LB120705
	Arsenic	20.0	+/-20.0	U	20.0	P	06/15/2022	17:40	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	17:40	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	17:40	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	17:40	LB120705
CCB05	Aluminum	100	+/-100	U	100	P	06/15/2022	18:34	LB120705
	Arsenic	20.0	+/-20.0	U	20.0	P	06/15/2022	18:34	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	18:34	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	18:34	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	18:34	LB120705
CCB06	Aluminum	100	+/-100	U	100	P	06/15/2022	19:12	LB120705
	Arsenic	20.0	+/-20.0	U	20.0	P	06/15/2022	19:12	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	19:12	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	19:12	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	19:12	LB120705
CCB07	Aluminum	100	+/-100	U	100	P	06/15/2022	20:02	LB120705
	Arsenic	20.0	+/-20.0	U	20.0	P	06/15/2022	20:02	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	20:02	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	20:02	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	20:02	LB120705
CCB08	Aluminum	100	+/-100	U	100	P	06/15/2022	20:48	LB120705
	Arsenic	20.0	+/-20.0	U	20.0	P	06/15/2022	20:48	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	20:48	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	20:48	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	20:48	LB120705



Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Inframark **SDG No.:** N3259
Contract: INFR01 **Lab Code:** CHEM **Case No.:** N3259 **SAS No.:** N3259

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Aluminum	100	+/-100	U	100	P	06/15/2022	21:34	LB120705
	Arsenic	20.0	+/-20.0	U	20.0	P	06/15/2022	21:34	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	21:34	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	21:34	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	21:34	LB120705
CCB10	Aluminum	100	+/-100	U	100	P	06/15/2022	22:01	LB120705
	Arsenic	6.86	+/-20.0	J	20.0	P	06/15/2022	22:01	LB120705
	Copper	20.0	+/-20.0	U	20.0	P	06/15/2022	22:01	LB120705
	Iron	100	+/-100	U	100	P	06/15/2022	22:01	LB120705
	Zinc	40.0	+/-40.0	U	40.0	P	06/15/2022	22:01	LB120705



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Inframark

SDG No.: N3259

Contract: INFR01

Lab Code: CHEM

Case No.: N3259

SAS No.: N3259

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------