

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Inframark **SDG No.:** N3259
Contract: INFR01 **Lab Code:** CHEM **Case No.:** N3259 **SAS No.:** N3259
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	237000	250000	95	210000	290000	06/10/2022	16:09	LB120624
	Arsenic	15.0			-20	20	06/10/2022	16:09	LB120624
	Copper	12.9	2.0	645	-18	22	06/10/2022	16:09	LB120624
	Iron	89100	100000	89	85000	120000	06/10/2022	16:09	LB120624
	Zinc	-0.80			-40	40	06/10/2022	16:09	LB120624
ICSAB01	Aluminum	263000	250000	105	210000	290000	06/10/2022	16:14	LB120624
	Arsenic	103	100	103	85	120	06/10/2022	16:14	LB120624
	Copper	525	510	103	430	590	06/10/2022	16:14	LB120624
	Iron	98800	99000	100	84000	110000	06/10/2022	16:14	LB120624
	Zinc	1020	950	107	810	1100	06/10/2022	16:14	LB120624
ICSA	Aluminum	249000	250000	100	210000	290000	06/10/2022	16:17	LB120624
	Arsenic	-51.0			-20	20	06/10/2022	16:17	LB120624
	Copper	11.0	2.0	550	-18	22	06/10/2022	16:17	LB120624
	Iron	103000	100000	103	85000	120000	06/10/2022	16:17	LB120624
	Zinc	-68.8			-40	40	06/10/2022	16:17	LB120624
ICSAB	Aluminum	244000	250000	98	210000	290000	06/10/2022	16:21	LB120624
	Arsenic	202	100	202	85	120	06/10/2022	16:21	LB120624
	Copper	531	510	104	430	590	06/10/2022	16:21	LB120624
	Iron	101000	99000	102	84000	110000	06/10/2022	16:21	LB120624
	Zinc	896	950	94	810	1100	06/10/2022	16:21	LB120624
ICSA01	Aluminum	263000	250000	105	210000	290000	06/14/2022	15:13	LB120678
	Arsenic	15.1			-20	20	06/14/2022	15:13	LB120678
	Copper	10.7	2.0	534	-18	22	06/14/2022	15:13	LB120678
	Iron	100000	100000	100	85000	120000	06/14/2022	15:13	LB120678
	Zinc	-5.21			-40	40	06/14/2022	15:13	LB120678
ICSAB01	Aluminum	258000	250000	103	210000	290000	06/14/2022	15:16	LB120678
	Arsenic	115	100	115	85	120	06/14/2022	15:16	LB120678
	Copper	512	510	100	430	590	06/14/2022	15:16	LB120678
	Iron	98700	99000	100	84000	110000	06/14/2022	15:16	LB120678
	Zinc	1020	950	107	810	1100	06/14/2022	15:16	LB120678
ICSA	Aluminum	249000	250000	100	210000	290000	06/14/2022	15:20	LB120678
	Arsenic	41.7			-20	20	06/14/2022	15:20	LB120678
	Copper	10.9	2.0	548	-18	22	06/14/2022	15:20	LB120678
	Iron	103000	100000	103	85000	120000	06/14/2022	15:20	LB120678
	Zinc	-129			-40	40	06/14/2022	15:20	LB120678
ICSAB	Aluminum	240000	250000	96	210000	290000	06/14/2022	15:24	LB120678
	Arsenic	143	100	143	85	120	06/14/2022	15:24	LB120678
	Copper	461	510	90	430	590	06/14/2022	15:24	LB120678
	Iron	99100	99000	100	84000	110000	06/14/2022	15:24	LB120678
	Zinc	868	950	91	810	1100	06/14/2022	15:24	LB120678
ICSA01	Aluminum	261000	250000	104	210000	290000	06/15/2022	14:51	LB120705
	Arsenic	12.5			-20	20	06/15/2022	14:51	LB120705

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Inframark SDG No.: N3259
Contract: INFR01 Lab Code: CHEM Case No.: N3259 SAS No.: N3259
ICS Source: EPA Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	16.4	2.0	820	-18	22	06/15/2022	14:51	LB120705
	Iron	101000	100000	101	85000	120000	06/15/2022	14:51	LB120705
	Zinc	-2.91			-40	40	06/15/2022	14:51	LB120705
ICSAB01	Aluminum	255000	250000	102	210000	290000	06/15/2022	14:55	LB120705
	Arsenic	114	100	114	85	120	06/15/2022	14:55	LB120705
	Copper	501	510	98	430	590	06/15/2022	14:55	LB120705
	Iron	98200	99000	99	84000	110000	06/15/2022	14:55	LB120705
	Zinc	1030	950	108	810	1100	06/15/2022	14:55	LB120705
ICSA	Aluminum	247000	250000	99	210000	290000	06/15/2022	14:59	LB120705
	Arsenic	20.1			-20	20	06/15/2022	14:59	LB120705
	Copper	65.4	2.0	3270	-18	22	06/15/2022	14:59	LB120705
	Iron	102000	100000	102	85000	120000	06/15/2022	14:59	LB120705
	Zinc	-75.5			-40	40	06/15/2022	14:59	LB120705
ICSAB	Aluminum	241000	250000	96	210000	290000	06/15/2022	15:02	LB120705
	Arsenic	173	100	173	85	120	06/15/2022	15:02	LB120705
	Copper	458	510	90	430	590	06/15/2022	15:02	LB120705
	Iron	99700	99000	101	84000	110000	06/15/2022	15:02	LB120705
	Zinc	930	950	98	810	1100	06/15/2022	15:02	LB120705