



metals
- 5a -
MATRIX SPIKE SUMMARY

client: Inframark level: low sdg no.: N3259
contract: INFR01 lab code: CHEM case no.: N3259 sas no.: N3259
matrix: Water sample id: N3252-01 client id: POOL-WATERMS
Percent Solids for Sample: NA Spiked ID: N3252-01MS Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	1040		54.1		1000	99		P
Arsenic	ug/L	75 - 125	396		10.0	U	400	99		P
Copper	ug/L	75 - 125	149		10.0	U	150	99		P
Iron	ug/L	75 - 125	1580		38.2	J	1500	103		P
Zinc	ug/L	75 - 125	167		65.0		100	102		P



metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Inframark level: low sdg no.: N3259
contract: INFR01 lab code: CHEM case no.: N3259 sas no.: N3259
matrix: Water sample id: N3252-01 client id: POOL-WATERMSD
Percent Solids for Sample: NA Spiked ID: N3252-01MSD Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	1060		54.1		1000	100		P
Arsenic	ug/L	75 - 125	394		10.0	U	400	98		P
Copper	ug/L	75 - 125	153		10.0	U	150	102		P
Iron	ug/L	75 - 125	1610		38.2	J	1500	105		P
Zinc	ug/L	75 - 125	168		65.0		100	103		P