



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Inframark **Level:** LOW **SDG No.:** N3259
Contract: INFR01 **Lab Code:** CHEM **Case No.:** N3259 **SAS No.:** N3259
Matrix: Water **Sample ID:** N3252-01 **Client ID:** POOL-WATERDUP
Percent Solids for Sample: NA **Duplicate ID** N3252-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	54.1		46.3	J	16		P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Copper	ug/L	20	10.0	U	10.0	U			P
Iron	ug/L	20	38.2	J	38.6	J	1		P
Zinc	ug/L	20	65.0		64.2		1		P



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Inframark **Level:** LOW **SDG No.:** N3259
Contract: INFR01 **Lab Code:** CHEM **Case No.:** N3259 **SAS No.:** N3259
Matrix: Water **Sample ID:** N3252-01MS **Client ID:** POOL-WATERMSD
Percent Solids for Sample: NA **Duplicate ID** N3252-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	1040		1060		2		P
Arsenic	ug/L	20	396		394		1		P
Copper	ug/L	20	149		153		3		P
Iron	ug/L	20	1580		1610		2		P
Zinc	ug/L	20	167		168		1		P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”