



- 7 -

Client:	<u>Inframark</u>	SDG No.:	<u>N3259</u>				
Contract:	INFR01	Lab Code:	CHEM	Case No.:	N3259	SAS No.:	N3259

Analyte	Units	True Value	Result	C	%	Acceptance	M
					Recovery	Limits	
PB145317BS							
Aluminum	ug/L	1000	929		93	85 - 115	P
Arsenic	ug/L	400	377		94	85 - 115	P
Copper	ug/L	150	142		95	85 - 115	P
Iron	ug/L	1500	1480		99	85 - 115	P
Zinc	ug/L	100	99.5		100	85 - 115	P