



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

Instrument ID: ECD_L Date(s) Analyzed: 06/06/2022 06/06/2022

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	7.04	6.94	7.14	25672300
		2	7.63	7.53	7.73	31484700
		3	8.39	8.29	8.49	105965000
		4	8.47	8.37	8.57	129973000
		5	9.37	9.27	9.47	22737900
Toxaphene	500	1	8.71	8.61	8.81	4074280
		2	9.03	8.93	9.13	4184000
		3	9.47	9.37	9.57	6095960
		4	9.56	9.46	9.66	11498700
		5	10.09	9.99	10.19	8882320



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

Instrument ID: ECD_L Date(s) Analyzed: 06/06/2022 06/06/2022

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	5.96	5.86	6.06	55658200
		2	6.64	6.54	6.74	66342900
		3	7.34	7.24	7.44	175372000
		4	7.41	7.31	7.51	186813000
		5	8.12	8.02	8.22	48348800
Toxaphene	500	1	7.72	7.62	7.82	9172520
		2	8.39	8.29	8.49	29087200
		3	8.81	8.71	8.91	14361800
		4	9.04	8.94	9.14	23250300
		5	9.18	9.08	9.28	32779400