



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

Continuing Calib Date: 06/09/2022 Initial Calibration Date(s): 06/06/2022 06/06/2022

Continuing Calib Time: 16:09 Initial Calibration Time(s): 14:25 15:31

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|-------|------------|
| Decachlorobiphenyl   | 11.64      | 11.64     | 11.54                  | 11.74 | 0.00       |
| Tetrachloro-m-xylene | 5.64       | 5.64      | 5.54                   | 5.74  | 0.00       |
| alpha-BHC            | 6.23       | 6.23      | 6.13                   | 6.33  | 0.00       |
| beta-BHC             | 6.88       | 6.88      | 6.78                   | 6.98  | 0.00       |
| delta-BHC            | 7.15       | 7.15      | 7.05                   | 7.25  | 0.00       |
| Heptachlor           | 7.28       | 7.28      | 7.18                   | 7.38  | 0.00       |
| 4,4'-DDE             | 8.66       | 8.66      | 8.56                   | 8.76  | 0.00       |
| Endrin               | 9.06       | 9.06      | 8.96                   | 9.16  | 0.00       |
| 4,4'-DDD             | 9.20       | 9.20      | 9.10                   | 9.30  | 0.00       |
| 4,4'-DDT             | 9.52       | 9.52      | 9.42                   | 9.62  | 0.00       |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

Continuing Calib Date: 06/09/2022 Initial Calibration Date(s): 06/06/2022 06/06/2022

Continuing Calib Time: 16:09 Initial Calibration Time(s): 14:25 15:31

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|-------|------------|
| Decachlorobiphenyl   | 10.38      | 10.38     | 10.28                  | 10.48 | 0.00       |
| Tetrachloro-m-xylene | 4.64       | 4.64      | 4.54                   | 4.74  | 0.00       |
| alpha-BHC            | 5.34       | 5.34      | 5.24                   | 5.44  | 0.00       |
| beta-BHC             | 6.14       | 6.14      | 6.04                   | 6.24  | 0.00       |
| delta-BHC            | 6.40       | 6.40      | 6.30                   | 6.50  | 0.00       |
| Heptachlor           | 6.17       | 6.17      | 6.07                   | 6.27  | 0.00       |
| 4,4'-DDE             | 7.63       | 7.62      | 7.52                   | 7.72  | -0.01      |
| Endrin               | 8.05       | 8.05      | 7.95                   | 8.15  | 0.00       |
| 4,4'-DDD             | 8.21       | 8.21      | 8.11                   | 8.31  | 0.00       |
| 4,4'-DDT             | 8.47       | 8.47      | 8.37                   | 8.57  | 0.00       |