



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

Continuing Calib Date: 06/09/2022 Initial Calibration Date(s): 06/06/2022 06/06/2022

Continuing Calib Time: 18:06 Initial Calibration Time(s): 14:25 15:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	11.64	11.64	11.54	11.74	0.00
Tetrachloro-m-xylene	5.64	5.64	5.54	5.74	0.00
alpha-BHC	6.23	6.23	6.13	6.33	0.01
beta-BHC	6.88	6.88	6.78	6.98	0.00
delta-BHC	7.15	7.15	7.05	7.25	0.00
Heptachlor	7.28	7.28	7.18	7.38	0.00
4,4'-DDE	8.66	8.66	8.56	8.76	0.00
Endrin	9.06	9.06	8.96	9.16	0.00
4,4'-DDD	9.20	9.20	9.10	9.30	0.00
4,4'-DDT	9.52	9.52	9.42	9.62	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

Continuing Calib Date: 06/09/2022 Initial Calibration Date(s): 06/06/2022 06/06/2022

Continuing Calib Time: 18:06 Initial Calibration Time(s): 14:25 15:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	10.38	10.38	10.28	10.48	0.00
Tetrachloro-m-xylene	4.64	4.64	4.54	4.74	0.00
alpha-BHC	5.34	5.34	5.24	5.44	0.00
beta-BHC	6.14	6.14	6.04	6.24	0.00
delta-BHC	6.40	6.40	6.30	6.50	0.00
Heptachlor	6.17	6.17	6.07	6.27	0.00
4,4'-DDE	7.62	7.62	7.52	7.72	0.00
Endrin	8.04	8.05	7.95	8.15	0.01
4,4'-DDD	8.21	8.21	8.11	8.31	0.00
4,4'-DDT	8.47	8.47	8.37	8.57	0.00