



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/06/2022 06/06/2022

Client Sample No.: CCAL01 Date Analyzed: 06/09/2022

Lab Sample No.: PSTDCCC050 Data File : PL075670.D Time Analyzed: 16:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	9.203	9.100	9.300	53.640	50.000	7.3
4,4'-DDE	8.662	8.558	8.758	52.700	50.000	5.4
4,4'-DDT	9.523	9.419	9.619	46.010	50.000	-8.0
alpha-BHC	6.229	6.125	6.325	54.680	50.000	9.4
beta-BHC	6.879	6.775	6.975	52.190	50.000	4.4
Decachlorobiphenyl	11.642	11.536	11.736	53.070	50.000	6.1
delta-BHC	7.149	7.045	7.245	58.480	50.000	17.0
Endrin	9.064	8.959	9.159	50.250	50.000	0.5
Heptachlor	7.282	7.179	7.379	50.930	50.000	1.9
Tetrachloro-m-xylene	5.644	5.541	5.741	53.070	50.000	6.1



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/06/2022 06/06/2022

Client Sample No.: CCAL01 Date Analyzed: 06/09/2022

Lab Sample No.: PSTDCCC050 Data File : PL075670.D Time Analyzed: 16:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	8.209	8.108	8.308	49.060	50.000	-1.9
4,4'-DDE	7.626	7.524	7.724	47.980	50.000	-4.0
4,4'-DDT	8.470	8.369	8.569	44.940	50.000	-10.1
alpha-BHC	5.340	5.240	5.440	51.050	50.000	2.1
beta-BHC	6.140	6.040	6.240	48.460	50.000	-3.1
Decachlorobiphenyl	10.384	10.283	10.483	41.790	50.000	-16.4
delta-BHC	6.400	6.300	6.500	49.290	50.000	-1.4
Endrin	8.045	7.945	8.145	50.550	50.000	1.1
Heptachlor	6.169	6.069	6.269	47.190	50.000	-5.6
Tetrachloro-m-xylene	4.637	4.538	4.738	49.150	50.000	-1.7