



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/06/2022 06/06/2022

Client Sample No.: CCAL02 Date Analyzed: 06/09/2022

Lab Sample No.: PSTDCCC050 Data File : PL075677.D Time Analyzed: 18:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	9.201	9.100	9.300	53.340	50.000	6.7
4,4'-DDE	8.659	8.558	8.758	52.740	50.000	5.5
4,4'-DDT	9.521	9.419	9.619	45.800	50.000	-8.4
alpha-BHC	6.225	6.125	6.325	54.050	50.000	8.1
beta-BHC	6.876	6.775	6.975	52.250	50.000	4.5
Decachlorobiphenyl	11.638	11.536	11.736	50.530	50.000	1.1
delta-BHC	7.147	7.045	7.245	58.070	50.000	16.1
Endrin	9.062	8.959	9.159	51.050	50.000	2.1
Heptachlor	7.279	7.179	7.379	51.100	50.000	2.2
Tetrachloro-m-xylene	5.641	5.541	5.741	52.320	50.000	4.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/06/2022 06/06/2022

Client Sample No.: CCAL02 Date Analyzed: 06/09/2022

Lab Sample No.: PSTDCCC050 Data File : PL075677.D Time Analyzed: 18:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	8.208	8.108	8.308	48.320	50.000	-3.4
4,4'-DDE	7.624	7.524	7.724	48.950	50.000	-2.1
4,4'-DDT	8.469	8.369	8.569	44.400	50.000	-11.2
alpha-BHC	5.339	5.240	5.440	50.910	50.000	1.8
beta-BHC	6.138	6.040	6.240	48.100	50.000	-3.8
Decachlorobiphenyl	10.383	10.283	10.483	41.940	50.000	-16.1
delta-BHC	6.399	6.300	6.500	49.070	50.000	-1.9
Endrin	8.044	7.945	8.145	51.570	50.000	3.1
Heptachlor	6.168	6.069	6.269	47.500	50.000	-5.0
Tetrachloro-m-xylene	4.636	4.538	4.738	47.130	50.000	-5.7