



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/06/2022 06/06/2022

Client Sample No.: CCAL03 Date Analyzed: 06/10/2022

Lab Sample No.: PSTDCCC050 Data File : PL075681.D Time Analyzed: 13:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	9.200	9.100	9.300	52.180	50.000	4.4
4,4'-DDE	8.658	8.558	8.758	51.480	50.000	3.0
4,4'-DDT	9.519	9.419	9.619	44.790	50.000	-10.4
alpha-BHC	6.225	6.125	6.325	51.340	50.000	2.7
beta-BHC	6.876	6.775	6.975	50.080	50.000	0.2
Decachlorobiphenyl	11.637	11.536	11.736	47.440	50.000	-5.1
delta-BHC	7.146	7.045	7.245	56.330	50.000	12.7
Endrin	9.060	8.959	9.159	48.480	50.000	-3.0
Heptachlor	7.279	7.179	7.379	48.560	50.000	-2.9
Tetrachloro-m-xylene	5.641	5.541	5.741	49.530	50.000	-0.9



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/06/2022 06/06/2022

Client Sample No.: CCAL03 Date Analyzed: 06/10/2022

Lab Sample No.: PSTDCCC050 Data File : PL075681.D Time Analyzed: 13:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	8.206	8.108	8.308	47.110	50.000	-5.8
4,4'-DDE	7.623	7.524	7.724	45.640	50.000	-8.7
4,4'-DDT	8.468	8.369	8.569	42.580	50.000	-14.8
alpha-BHC	5.339	5.240	5.440	46.540	50.000	-6.9
beta-BHC	6.138	6.040	6.240	43.650	50.000	-12.7
Decachlorobiphenyl	10.382	10.283	10.483	40.880	50.000	-18.2
delta-BHC	6.399	6.300	6.500	45.300	50.000	-9.4
Endrin	8.043	7.945	8.145	47.090	50.000	-5.8
Heptachlor	6.168	6.069	6.269	43.840	50.000	-12.3
Tetrachloro-m-xylene	4.637	4.538	4.738	44.460	50.000	-11.1