



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/06/2022 06/06/2022

Client Sample No.: CCAL04 Date Analyzed: 06/10/2022

Lab Sample No.: PSTDCCC050 Data File : PL075695.D Time Analyzed: 19:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	9.200	9.100	9.300	51.850	50.000	3.7
4,4'-DDE	8.658	8.558	8.758	50.110	50.000	0.2
4,4'-DDT	9.519	9.419	9.619	42.650	50.000	-14.7
alpha-BHC	6.224	6.125	6.325	49.790	50.000	-0.4
beta-BHC	6.875	6.775	6.975	48.780	50.000	-2.4
Decachlorobiphenyl	11.637	11.536	11.736	45.440	50.000	-9.1
delta-BHC	7.145	7.045	7.245	54.790	50.000	9.6
Endrin	9.060	8.959	9.159	47.560	50.000	-4.9
Heptachlor	7.278	7.179	7.379	47.160	50.000	-5.7
Tetrachloro-m-xylene	5.641	5.541	5.741	48.250	50.000	-3.5



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: INFR01

Lab Code: CHEM Case No.: N3259 SAS No.: N3259 SDG NO.: N3259

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/06/2022 06/06/2022

Client Sample No.: CCAL04 Date Analyzed: 06/10/2022

Lab Sample No.: PSTDCCC050 Data File : PL075695.D Time Analyzed: 19:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	8.204	8.108	8.308	48.480	50.000	-3.0
4,4'-DDE	7.622	7.524	7.724	46.730	50.000	-6.5
4,4'-DDT	8.466	8.369	8.569	42.180	50.000	-15.6
alpha-BHC	5.337	5.240	5.440	48.000	50.000	-4.0
beta-BHC	6.137	6.040	6.240	44.680	50.000	-10.6
Decachlorobiphenyl	10.381	10.283	10.483	41.780	50.000	-16.4
delta-BHC	6.398	6.300	6.500	46.350	50.000	-7.3
Endrin	8.042	7.945	8.145	48.000	50.000	-4.0
Heptachlor	6.166	6.069	6.269	45.390	50.000	-9.2
Tetrachloro-m-xylene	4.635	4.538	4.738	45.100	50.000	-9.8