



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL060622

Review By	Abdul	Review On	6/7/2022 8:49:39 AM				
Supervise By	Ankita	Supervise On	6/7/2022 3:40:02 PM				
SubDirectory	PL060622	HP Acquire Method	HP Processing Method	PL060622			
STD. NAME		STD REF.#					
Tune/Reschk		PP19605,P10884					
Initial Calibration Stds		PP19748,PP19749,PP19750,PP19751,PP19752,PP19753,PP19754,PP19755,PP19756,PP19757,PP19758,PP19759,PP19760,PP19761,PP19762					
CCC		PP19750,PP19755,PP19760					
Internal Standard/PEM							
ICV/I.BLK		PP19764					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL075556.D	06 Jun 2022 13:19		AR\AJ	Ok
2	I.BLK	I.BLK	PL075557.D	06 Jun 2022 13:36		AR\AJ	Ok
3	PEM	PEM	PL075558.D	06 Jun 2022 13:52		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL075559.D	06 Jun 2022 14:09		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PL075560.D	06 Jun 2022 14:25		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL075561.D	06 Jun 2022 14:42		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL075562.D	06 Jun 2022 14:58		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL075563.D	06 Jun 2022 15:15		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL075564.D	06 Jun 2022 15:31		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL075565.D	06 Jun 2022 15:48		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL075566.D	06 Jun 2022 16:04		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL075567.D	06 Jun 2022 16:21		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL075568.D	06 Jun 2022 16:37		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL075569.D	06 Jun 2022 16:54		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PL075570.D	06 Jun 2022 17:10		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL075571.D	06 Jun 2022 17:27		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL075572.D	06 Jun 2022 17:43		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL075573.D	06 Jun 2022 18:00		AR\AJ	Ok,M
19	PTOXICC100	PTOXICC100	PL075574.D	06 Jun 2022 18:17		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL060622

Review By	Abdul	Review On	6/7/2022 8:49:39 AM
Supervise By	Ankita	Supervise On	6/7/2022 3:40:02 PM
SubDirectory	PL060622	HP Acquire Method	HP Processing Method PL060622
STD. NAME	STD REF.#		
Tune/Reschk	PP19605,PP10884		
Initial Calibration Stds	PP19748,PP19749,PP19750,PP19751,PP19752,PP19753,PP19754,PP19755,PP19756,PP19757,PP19758,PP19759,PP19760,PP19761,PP19762		
CCC	PP19750,PP19755,PP19760		
Internal Standard/PEM			
ICV/I.BLK	PP19764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			
20	PSTDICV050	ICVPL060622	PL075575.D 06 Jun 2022 19:04 Not used AR\AJ Not Ok
21	PCHLORICV500	ICVPL060622CHLOR	PL075576.D 06 Jun 2022 19:21 AR\AJ Ok
22	PTOXICV500	ICVPL060622TOX	PL075577.D 06 Jun 2022 19:37 AR\AJ Ok
23	PSTDICV050	ICVPL060622	PL075578.D 06 Jun 2022 20:27 AR\AJ Ok
24	N2656-21	PT-PEST-SOIL	PL075579.D 06 Jun 2022 22:06 Need 10X AR\AJ Dilution
25	N2656-21DL	PT-PEST-SOILDL	PL075580.D 06 Jun 2022 22:22 Need more dilution AR\AJ Not Ok
26	N2656-21DL	PT-PEST-SOILDL	PL075581.D 06 Jun 2022 22:39 Need 20X AR\AJ Dilution
27	N2656-21DL2	PT-PEST-SOILDL2	PL075582.D 06 Jun 2022 22:56 AR\AJ Ok
28	I.BLK	I.BLK	PL075583.D 07 Jun 2022 01:08 AR\AJ Ok
29	PEM	PEM	PL075584.D 07 Jun 2022 01:24 AR\AJ Ok,M
30	PSTDCCC050	PSTDCCC050	PL075585.D 07 Jun 2022 01:41 AR\AJ Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL060922

Review By	Abdul	Review On	6/10/2022 8:39:44 AM				
Supervise By	Ankita	Supervise On	6/10/2022 9:29:12 AM				
SubDirectory	PL060922	HP Acquire Method	HP Processing Method				
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		PP19748,PP19749,PP19750,PP19751,PP19752,PP19753,PP19754,PP19755,PP19756,PP19757,PP19758,PP19759,PP19760,PP19761,PP19762					
CCC		PP19750,PP19755,PP19760					
Internal Standard/PEM							
ICV/I.BLK		PP19764					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL075656.D	09 Jun 2022 08:56		AR\AJ	Ok
2	I.BLK	I.BLK	PL075657.D	09 Jun 2022 09:48		AR\AJ	Ok
3	PEM	PEM	PL075658.D	09 Jun 2022 10:06		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL075659.D	09 Jun 2022 11:19		AR\AJ	Ok
5	PB145427BL	PB145427BL	PL075660.D	09 Jun 2022 12:34		AR\AJ	Ok
6	PB145427BS	PB145427BS	PL075661.D	09 Jun 2022 12:51	rec fail	AR\AJ	Not Ok
7	N3233-01	CR-1	PL075662.D	09 Jun 2022 13:07		AR\AJ	Ok
8	N3233-01MS	CR-1MS	PL075663.D	09 Jun 2022 13:24		AR\AJ	Ok
9	N3233-01MSD	CR-1MSD	PL075664.D	09 Jun 2022 13:40		AR\AJ	Ok
10	N3233-02	CR-2	PL075665.D	09 Jun 2022 13:57		AR\AJ	Ok,M
11	N3234-02	RB-21499	PL075666.D	09 Jun 2022 14:13		AR\AJ	Ok,M
12	N3251-01	NB-8-060822	PL075667.D	09 Jun 2022 14:30		AR\AJ	Ok
13	N3267-01	OR-2-060822	PL075668.D	09 Jun 2022 14:46		AR\AJ	Ok,M
14	I.BLK	I.BLK	PL075669.D	09 Jun 2022 15:03		AR\AJ	Ok
15	PSTDCCC050	PSTDCCC050	PL075670.D	09 Jun 2022 16:09		AR\AJ	Ok
16	N3259-01	BI-WEEKLY-INFLUENT	PL075671.D	09 Jun 2022 16:27	DCB low in both column	AR\AJ	ReRun
17	N3261-01	BI-WEEKLY-DECANT	PL075672.D	09 Jun 2022 16:44	TCMX low in 2nd column & DCB low in 1st column	AR\AJ	ReRun
18	PB145434BS	PB145434BS	PL075673.D	09 Jun 2022 17:00	Comp # 7 having low rec	AR\AJ	Not Ok
19	PB145434BSD	PB145434BSD	PL075674.D	09 Jun 2022 17:17	Comp # 7 having low rec	AR\AJ	Not Ok



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL060922

Review By	Abdul	Review On	6/10/2022 8:39:44 AM
Supervise By	Ankita	Supervise On	6/10/2022 9:29:12 AM
SubDirectory	PL060922	HP Acquire Method	HP Processing Method
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP19748,PP19749,PP19750,PP19751,PP19752,PP19753,PP19754,PP19755,PP19756,PP19757,PP19758,PP19759,PP19760,PP19761,PP19762		
CCC	PP19750,PP19755,PP19760		
Internal Standard/PEM			
ICV/I.BLK	PP19764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

20	PB145434BL	PB145434BL	PL075675.D	09 Jun 2022 17:33		AR\AJ	Ok,M
21	I.BLK	I.BLK	PL075676.D	09 Jun 2022 17:50		AR\AJ	Ok
22	PSTDCCC050	PSTDCCC050	PL075677.D	09 Jun 2022 18:06		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL061022

Review By	Abdul	Review On	6/13/2022 9:57:33 AM				
Supervise By	Ankita	Supervise On	6/13/2022 11:36:19 AM				
SubDirectory	PL061022	HP Acquire Method	HP Processing Method				
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		PP19748,PP19749,PP19750,PP19751,PP19752,PP19753,PP19754,PP19755,PP19756,PP19757,PP19758,PP19759,PP19760,PP19761,PP19762					
CCC Internal Standard/PEM		PP19750,PP19755,PP19760					
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		PP19764					

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL075678.D	10 Jun 2022 08:47		AR\AJ	Ok
2	I.BLK	I.BLK	PL075679.D	10 Jun 2022 09:59		AR\AJ	Ok
3	PEM	PEM	PL075680.D	10 Jun 2022 10:15		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL075681.D	10 Jun 2022 13:42		AR\AJ	Ok,M
5	PB145370BS	PB145370BS	PL075682.D	10 Jun 2022 13:59		AR\AJ	Ok
6	N3296-01	OK-03-060922	PL075683.D	10 Jun 2022 14:59		AR\AJ	Ok,M
7	N3259-01RE	BI-WEEKLY-INFLUENT	PL075684.D	10 Jun 2022 15:15	DCB low in both column	AR\AJ	Confirm
8	N3261-01	BI-WEEKLY-DECANT	PL075685.D	10 Jun 2022 15:37		AR\AJ	Ok,M
9	PB145434BS	PB145434BS	PL075686.D	10 Jun 2022 17:28	Comp # 22 having high rec.	AR\AJ	Ok,M
10	PB145434BSD	PB145434BSD	PL075687.D	10 Jun 2022 17:45	Comp # 22 having high rec.	AR\AJ	Ok,M
11	PB145465BS	PB145465BS	PL075688.D	10 Jun 2022 18:01	Comp # 22 having high rec.	AR\AJ	Ok,M
12	PB145465BSD	PB145465BSD	PL075689.D	10 Jun 2022 18:18	Comp # 22 having high rec.	AR\AJ	Ok,M
13	PB145465BL	PB145465BL	PL075690.D	10 Jun 2022 18:34		AR\AJ	Ok
14	PB145452BL	PB145452BL	PL075691.D	10 Jun 2022 18:51		AR\AJ	Ok
15	PB145452BS	PB145452BS	PL075692.D	10 Jun 2022 19:07	All comp coming f flag in 1st column	AR\AJ	Not Ok
16	I.BLK	I.BLK	PL075693.D	10 Jun 2022 19:24		AR\AJ	Ok
17	PEM	PEM	PL075694.D	10 Jun 2022 19:40		AR\AJ	Ok,M
18	PSTDCCC050	PSTDCCC050	PL075695.D	10 Jun 2022 19:57		AR\AJ	Ok,M
19	N3299-03	SC-15	PL075696.D	10 Jun 2022 21:20		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL061022

Review By	Abdul	Review On	6/13/2022 9:57:33 AM				
Supervise By	Ankita	Supervise On	6/13/2022 11:36:19 AM				
SubDirectory	PL061022	HP Acquire Method	HP Processing Method				
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		PP19748,PP19749,PP19750,PP19751,PP19752,PP19753,PP19754,PP19755,PP19756,PP19757,PP19758,PP19759,PP19760,PP19761,PP19762					
CCC		PP19750,PP19755,PP19760					
Internal Standard/PEM							
ICV/I.BLK		PP19764					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

20	N3299-04	SC-16	PL075697.D	10 Jun 2022 21:36	Need 10x	AR\AJ	Dilution
21	N3299-05	SC-17	PL075698.D	10 Jun 2022 21:53	Need 10x	AR\AJ	Dilution
22	N3299-06	SC-18	PL075699.D	10 Jun 2022 22:09		AR\AJ	Ok
23	N3302-01	I-1	PL075700.D	10 Jun 2022 22:26		AR\AJ	Ok,M
24	N3302-02	I-2	PL075701.D	10 Jun 2022 22:42		AR\AJ	Ok,M
25	N3281-01	129	PL075702.D	10 Jun 2022 22:59		AR\AJ	Ok
26	N3282-01	125	PL075703.D	10 Jun 2022 23:16	TCMX & DCB coming f flag in 1st column	AR\AJ	Not Ok
27	N3297-02	TI-633	PL075704.D	10 Jun 2022 23:32		AR\AJ	Ok,M
28	N3297-04	TI-633-X	PL075705.D	10 Jun 2022 23:49	TCMX low in 2nd column,	AR\AJ	Ok,M
29	N3297-05	TI-633-EB	PL075706.D	11 Jun 2022 00:05		AR\AJ	Ok
30	I.BLK	I.BLK	PL075707.D	11 Jun 2022 00:22		AR\AJ	Ok
31	PSTDCCC050	PSTDCCC050	PL075708.D	11 Jun 2022 00:55		AR\AJ	Ok
32	N3295-01	OR-03-060922	PL075709.D	11 Jun 2022 02:34		AR\AJ	Ok,M
33	N3295-01MS	OR-03-060922MS	PL075710.D	11 Jun 2022 02:50	DCB high in 1st column, some comp. having high rec	AR\AJ	Ok,M
34	N3295-01MSD	OR-03-060922MSD	PL075711.D	11 Jun 2022 03:07	DCB high in 1st column, comp # 13 having high rec	AR\AJ	Ok,M
35	I.BLK	I.BLK	PL075712.D	11 Jun 2022 04:13		AR\AJ	Ok
36	PSTDCCC050	PSTDCCC050	PL075713.D	11 Jun 2022 04:46		AR\AJ	Ok

M : Manual Integration