



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: N3259

Client: Inframark

Analytical Method: 608.3 Pest

Datafile : PL075686.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB145434BS	alpha-BHC	0.0125	0.0089	ug/L	71				37	140	
	Heptachlor	0.0125	0.0096	ug/L	77				34	140	
	beta-BHC	0.0125	0.0094	ug/L	75				17	147	
	delta-BHC	0.0125	0.0057	ug/L	46				19	140	
	4,4'-DDE	0.0125	0.0096	ug/L	77				30	145	
	Endrin	0.0125	0.0096	ug/L	77				30	147	
	4,4'-DDD	0.0125	0.0096	ug/L	77				31	141	
	4,4'-DDT	0.0125	0.0090	ug/L	72				25	160	