



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: N3259

Client: Inframark

Analytical Method: 608.3 Pest

Datafile : PL075687.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB145434BSD	alpha-BHC	0.0125	0.0099	ug/L	79	11			37	140	20
	Heptachlor	0.0125	0.0097	ug/L	78	1			34	140	20
	beta-BHC	0.0125	0.0095	ug/L	76	1			17	147	20
	delta-BHC	0.0125	0.0041	ug/L	33	33		*	19	140	20
	4,4'-DDE	0.0125	0.0098	ug/L	78	1			30	145	20
	Endrin	0.0125	0.011	ug/L	86	11			30	147	20
	4,4'-DDD	0.0125	0.0097	ug/L	78	1			31	141	20
	4,4'-DDT	0.0125	0.0084	ug/L	67	7			25	160	20