



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

Continuing Calib Date: 12/13/2022 Initial Calibration Date(s): 11/25/2022 11/26/2022

Continuing Calib Time: 13:22 Initial Calibration Time(s): 20:41 04:21

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.47	5.48	5.38	5.58	0.01
Aroclor-1016-2	(2)	5.50	5.50	5.40	5.60	0.01
Aroclor-1016-3	(3)	5.56	5.56	5.46	5.66	0.00
Aroclor-1016-4	(4)	5.66	5.66	5.56	5.76	0.00
Aroclor-1016-5	(5)	5.95	5.96	5.86	6.06	0.01
Aroclor-1260-1	(1)	7.09	7.09	6.99	7.19	0.00
Aroclor-1260-2	(2)	7.35	7.35	7.25	7.45	0.00
Aroclor-1260-3	(3)	7.71	7.71	7.61	7.81	0.00
Aroclor-1260-4	(4)	7.94	7.94	7.84	8.04	0.01
Aroclor-1260-5	(5)	8.25	8.26	8.16	8.36	0.01
Tetrachloro-m-xylene		4.30	4.30	4.20	4.40	0.00
Decachlorobiphenyl		10.04	10.05	9.95	10.15	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

Continuing Calib Date: 12/13/2022 Initial Calibration Date(s): 11/25/2022 11/26/2022

Continuing Calib Time: 13:22 Initial Calibration Time(s): 20:41 04:21

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1 (1)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-2 (2)	4.58	4.58	4.48	4.68	0.00
Aroclor-1016-3 (3)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-4 (4)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-5 (5)	5.01	5.01	4.91	5.11	0.00
Aroclor-1260-1 (1)	6.05	6.05	5.95	6.15	0.00
Aroclor-1260-2 (2)	6.24	6.24	6.14	6.34	0.00
Aroclor-1260-3 (3)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-4 (4)	6.86	6.87	6.77	6.97	0.01
Aroclor-1260-5 (5)	7.11	7.11	7.01	7.21	0.00
Tetrachloro-m-xylene	3.47	3.48	3.38	3.58	0.01
Decachlorobiphenyl	8.47	8.48	8.38	8.58	0.01