



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

Continuing Calib Date: 12/13/2022 Initial Calibration Date(s): 11/25/2022 11/26/2022

Continuing Calib Time: 13:39 Initial Calibration Time(s): 20:41 04:21

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1242-1	(1)	5.47	5.48	5.38	5.58	0.01
Aroclor-1242-2	(2)	5.49	5.50	5.40	5.60	0.01
Aroclor-1242-3	(3)	5.56	5.56	5.46	5.66	0.00
Aroclor-1242-4	(4)	5.66	5.66	5.56	5.76	0.00
Aroclor-1242-5	(5)	6.40	6.40	6.30	6.50	0.00
Tetrachloro-m-xylene		4.30	4.30	4.20	4.40	0.00
Decachlorobiphenyl		10.04	10.05	9.95	10.15	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

Continuing Calib Date: 12/13/2022 Initial Calibration Date(s): 11/25/2022 11/26/2022

Continuing Calib Time: 13:39 Initial Calibration Time(s): 20:41 04:21

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.56	4.56	4.46	4.66	0.00
Aroclor-1242-2 (2)	4.58	4.58	4.48	4.68	0.00
Aroclor-1242-3 (3)	4.75	4.76	4.66	4.86	0.01
Aroclor-1242-4 (4)	4.84	4.84	4.74	4.94	0.00
Aroclor-1242-5 (5)	5.36	5.37	5.27	5.47	0.01
Tetrachloro-m-xylene	3.47	3.48	3.38	3.58	0.01
Decachlorobiphenyl	8.47	8.48	8.38	8.58	0.01