



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

Continuing Calib Date: 12/13/2022 Initial Calibration Date(s): 11/25/2022 11/26/2022

Continuing Calib Time: 14:13 Initial Calibration Time(s): 20:41 04:21

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1254-2 (2)	6.55	6.56	6.46	6.66	0.01
Aroclor-1254-3 (3)	6.92	6.93	6.83	7.03	0.01
Aroclor-1254-4 (4)	7.21	7.21	7.11	7.31	0.00
Aroclor-1254-5 (5)	7.63	7.64	7.54	7.74	0.01
Tetrachloro-m-xylene	4.30	4.30	4.20	4.40	0.00
Decachlorobiphenyl	10.04	10.05	9.95	10.15	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

Continuing Calib Date: 12/13/2022 Initial Calibration Date(s): 11/25/2022 11/26/2022

Continuing Calib Time: 14:13 Initial Calibration Time(s): 20:41 04:21

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.36	5.37	5.27	5.47	0.01
Aroclor-1254-2 (2)	5.51	5.52	5.42	5.62	0.01
Aroclor-1254-3 (3)	5.91	5.92	5.82	6.02	0.01
Aroclor-1254-4 (4)	6.15	6.15	6.05	6.25	0.01
Aroclor-1254-5 (5)	6.56	6.57	6.47	6.67	0.01
Tetrachloro-m-xylene	3.47	3.48	3.38	3.58	0.01
Decachlorobiphenyl	8.47	8.48	8.38	8.58	0.01