



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

Continuing Calib Date: 12/13/2022 Initial Calibration Date(s): 11/25/2022 11/26/2022

Continuing Calib Time: 22:44 Initial Calibration Time(s): 20:41 04:21

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1248-1 (1)	5.47	5.48	5.38	5.58	0.01
Aroclor-1248-2 (2)	5.75	5.75	5.65	5.85	0.00
Aroclor-1248-3 (3)	5.95	5.96	5.86	6.06	0.01
Aroclor-1248-4 (4)	6.36	6.36	6.26	6.46	0.00
Aroclor-1248-5 (5)	6.40	6.40	6.30	6.50	0.00
Tetrachloro-m-xylene	4.30	4.30	4.20	4.40	0.00
Decachlorobiphenyl	10.04	10.05	9.95	10.15	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

Continuing Calib Date: 12/13/2022 Initial Calibration Date(s): 11/25/2022 11/26/2022

Continuing Calib Time: 22:44 Initial Calibration Time(s): 20:41 04:21

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1248-1	(1)	4.56	4.56	4.46	4.66	0.00
Aroclor-1248-2	(2)	4.79	4.80	4.70	4.90	0.01
Aroclor-1248-3	(3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1248-4	(4)	5.01	5.01	4.91	5.11	0.00
Aroclor-1248-5	(5)	5.40	5.41	5.31	5.51	0.01
Tetrachloro-m-xylene		3.47	3.48	3.38	3.58	0.01
Decachlorobiphenyl		8.47	8.48	8.38	8.58	0.01