



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2022 11/25/2022

Client Sample No.: CCAL01 Date Analyzed: 12/13/2022

Lab Sample No.: AR1660CCC500 Data File : PO091297.D Time Analyzed: 13:22

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                      |        | FROM      | TO     |                    |                   |       |
| Aroclor-1016-1       | 5.473  | 5.375     | 5.575  | 460.160            | 500.000           | -8.0  |
| Aroclor-1016-2       | 5.495  | 5.397     | 5.597  | 455.230            | 500.000           | -9.0  |
| Aroclor-1016-3       | 5.557  | 5.459     | 5.659  | 445.330            | 500.000           | -10.9 |
| Aroclor-1016-4       | 5.656  | 5.558     | 5.758  | 454.990            | 500.000           | -9.0  |
| Aroclor-1016-5       | 5.954  | 5.856     | 6.056  | 436.880            | 500.000           | -12.6 |
| Aroclor-1260-1       | 7.088  | 6.992     | 7.192  | 446.110            | 500.000           | -10.8 |
| Aroclor-1260-2       | 7.348  | 7.251     | 7.451  | 442.470            | 500.000           | -11.5 |
| Aroclor-1260-3       | 7.709  | 7.613     | 7.813  | 456.280            | 500.000           | -8.7  |
| Aroclor-1260-4       | 7.935  | 7.839     | 8.039  | 468.290            | 500.000           | -6.3  |
| Aroclor-1260-5       | 8.251  | 8.156     | 8.356  | 436.890            | 500.000           | -12.6 |
| Decachlorobiphenyl   | 10.044 | 9.950     | 10.150 | 42.680             | 50.000            | -14.6 |
| Tetrachloro-m-xylene | 4.296  | 4.198     | 4.398  | 46.090             | 50.000            | -7.8  |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2022 11/25/2022

Client Sample No.: CCAL01 Date Analyzed: 12/13/2022

Lab Sample No.: AR1660CCC500 Data File : PO091297.D Time Analyzed: 13:22

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.558 | 4.463     | 4.663 | 492.800            | 500.000           | -1.4 |
| Aroclor-1016-2       | 4.576 | 4.481     | 4.681 | 500.470            | 500.000           | 0.1  |
| Aroclor-1016-3       | 4.752 | 4.657     | 4.857 | 508.450            | 500.000           | 1.7  |
| Aroclor-1016-4       | 4.795 | 4.700     | 4.900 | 483.410            | 500.000           | -3.3 |
| Aroclor-1016-5       | 5.008 | 4.913     | 5.113 | 525.870            | 500.000           | 5.2  |
| Aroclor-1260-1       | 6.045 | 5.952     | 6.152 | 496.260            | 500.000           | -0.7 |
| Aroclor-1260-2       | 6.235 | 6.142     | 6.342 | 497.040            | 500.000           | -0.6 |
| Aroclor-1260-3       | 6.388 | 6.294     | 6.494 | 477.720            | 500.000           | -4.5 |
| Aroclor-1260-4       | 6.862 | 6.769     | 6.969 | 492.230            | 500.000           | -1.6 |
| Aroclor-1260-5       | 7.106 | 7.014     | 7.214 | 487.610            | 500.000           | -2.5 |
| Decachlorobiphenyl   | 8.473 | 8.382     | 8.582 | 48.800             | 50.000            | -2.4 |
| Tetrachloro-m-xylene | 3.473 | 3.376     | 3.576 | 48.570             | 50.000            | -2.9 |