



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2022 11/25/2022

Client Sample No.: CCAL02 Date Analyzed: 12/13/2022

Lab Sample No.: AR1660CCC500 Data File : PO091312.D Time Analyzed: 18:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.472	5.375	5.575	454.380	500.000	-9.1
Aroclor-1016-2	5.493	5.397	5.597	448.050	500.000	-10.4
Aroclor-1016-3	5.556	5.459	5.659	439.690	500.000	-12.1
Aroclor-1016-4	5.654	5.558	5.758	441.950	500.000	-11.6
Aroclor-1016-5	5.952	5.856	6.056	430.230	500.000	-14.0
Aroclor-1260-1	7.087	6.992	7.192	433.100	500.000	-13.4
Aroclor-1260-2	7.345	7.251	7.451	432.720	500.000	-13.5
Aroclor-1260-3	7.707	7.613	7.813	441.720	500.000	-11.7
Aroclor-1260-4	7.933	7.839	8.039	441.060	500.000	-11.8
Aroclor-1260-5	8.249	8.156	8.356	419.310	500.000	-16.1
Decachlorobiphenyl	10.039	9.950	10.150	41.570	50.000	-16.9
Tetrachloro-m-xylene	4.295	4.198	4.398	47.520	50.000	-5.0



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2022 11/25/2022

Client Sample No.: CCAL02 Date Analyzed: 12/13/2022

Lab Sample No.: AR1660CCC500 Data File : PO091312.D Time Analyzed: 18:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.557	4.463	4.663	516.810	500.000	3.4
Aroclor-1016-2	4.575	4.481	4.681	524.800	500.000	5.0
Aroclor-1016-3	4.750	4.657	4.857	524.390	500.000	4.9
Aroclor-1016-4	4.794	4.700	4.900	488.750	500.000	-2.3
Aroclor-1016-5	5.008	4.913	5.113	555.270	500.000	11.1
Aroclor-1260-1	6.044	5.952	6.152	509.510	500.000	1.9
Aroclor-1260-2	6.234	6.142	6.342	505.530	500.000	1.1
Aroclor-1260-3	6.387	6.294	6.494	474.880	500.000	-5.0
Aroclor-1260-4	6.861	6.769	6.969	494.130	500.000	-1.2
Aroclor-1260-5	7.106	7.014	7.214	479.690	500.000	-4.1
Decachlorobiphenyl	8.472	8.382	8.582	48.500	50.000	-3.0
Tetrachloro-m-xylene	3.473	3.376	3.576	50.990	50.000	2.0