



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2022 11/25/2022

Client Sample No.: CCAL03 Date Analyzed: 12/13/2022

Lab Sample No.: AR1660CCC500 Data File : PO091323.D Time Analyzed: 22:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.472	5.375	5.575	480.070	500.000	-4.0
Aroclor-1016-2	5.494	5.397	5.597	475.880	500.000	-4.8
Aroclor-1016-3	5.556	5.459	5.659	463.140	500.000	-7.4
Aroclor-1016-4	5.655	5.558	5.758	469.650	500.000	-6.1
Aroclor-1016-5	5.952	5.856	6.056	466.090	500.000	-6.8
Aroclor-1260-1	7.086	6.992	7.192	460.740	500.000	-7.9
Aroclor-1260-2	7.345	7.251	7.451	453.790	500.000	-9.2
Aroclor-1260-3	7.708	7.613	7.813	467.550	500.000	-6.5
Aroclor-1260-4	7.934	7.839	8.039	473.220	500.000	-5.4
Aroclor-1260-5	8.251	8.156	8.356	435.810	500.000	-12.8
Decachlorobiphenyl	10.041	9.950	10.150	43.910	50.000	-12.2
Tetrachloro-m-xylene	4.295	4.198	4.398	48.180	50.000	-3.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: N5992 SAS No.: N5992 SDG NO.: N5992

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2022 11/25/2022

Client Sample No.: CCAL03 Date Analyzed: 12/13/2022

Lab Sample No.: AR1660CCC500 Data File : PO091323.D Time Analyzed: 22:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.556	4.463	4.663	520.980	500.000	4.2
Aroclor-1016-2	4.575	4.481	4.681	521.820	500.000	4.4
Aroclor-1016-3	4.750	4.657	4.857	524.140	500.000	4.8
Aroclor-1016-4	4.793	4.700	4.900	487.320	500.000	-2.5
Aroclor-1016-5	5.007	4.913	5.113	558.140	500.000	11.6
Aroclor-1260-1	6.044	5.952	6.152	527.910	500.000	5.6
Aroclor-1260-2	6.235	6.142	6.342	528.000	500.000	5.6
Aroclor-1260-3	6.387	6.294	6.494	500.320	500.000	0.1
Aroclor-1260-4	6.861	6.769	6.969	507.010	500.000	1.4
Aroclor-1260-5	7.105	7.014	7.214	503.150	500.000	0.6
Decachlorobiphenyl	8.472	8.382	8.582	50.510	50.000	1.0
Tetrachloro-m-xylene	3.472	3.376	3.576	50.710	50.000	1.4