

Report of Analysis

Client:	Remington & Vernick Engineers	Date Collected:	12/12/22
Project:	Wildwood DPW Storage Yard Site - # 0514T294	Date Received:	12/12/22
Client Sample ID:	TP-QMS	SDG No.:	N5992
Lab Sample ID:	N6024-01MS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	88.3
Sample Wt/Vol:	30.09	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO091317.D	1	12/13/22 09:15	12/13/22 19:59	PB149582

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	181		3.40	19.2	ug/kg
11104-28-2	Aroclor-1221	5.30	U	5.30	19.2	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	19.2	ug/kg
53469-21-9	Aroclor-1242	2.70	U	2.70	19.2	ug/kg
12672-29-6	Aroclor-1248	3.40	U	3.40	19.2	ug/kg
11097-69-1	Aroclor-1254	4.80	U	4.80	19.2	ug/kg
37324-23-5	Aroclor-1262	3.80	U	3.80	19.2	ug/kg
11100-14-4	Aroclor-1268	6.50	U	6.50	19.2	ug/kg
11096-82-5	Aroclor-1260	162		3.70	19.2	ug/kg
Total PCBs	Total PCBs	343		7.10	19.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.4		40 - 162	107%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.1		32 - 176	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit