



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: N5992

Client: Remington & Vernick Engineers

Analytical Method: 8082A

DataFile : PO091317.D

Lab Sample ID:		Parameter	Spike	Sample Result	Result	Units	Rec	Qual	RPD	Qual	Low	Limits High	RPD
Client Sample ID:		TP-QMS											
N6024-01MS		AR1016	188.2	0	181	ug/kg	96				70	142	
		AR1260	188.2	0	162	ug/kg	86				50	147	