



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: N6070 SAS No.: N6070 SDG NO.: N6070

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/19/2022

Lab File ID: BN023280.D Time Analyzed: 11:22

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7661	7.999	24904	10.82	15496	14.65
UPPER LIMIT	15322	8.499	49808	11.319	30992	15.145
LOWER LIMIT	3830.5	7.499	12452	10.319	7748	14.145
EPA SAMPLE NO.						
01 PB149692BL	8904	8.00	28123	10.82	17435	14.65
02 GW-BR-04-226-245-121422	6913	8.00	22753	10.82	14155	14.65
03 OWBR-01-160-180-121422	7610	8.01	24602	10.82	15939	14.65
04 OWBR-02-160-180-121422	7230	8.00	23777	10.82	14567	14.65
05 OWBR-03-128-148-121422	7020	8.00	22963	10.82	14454	14.65
06 PB149692BS	8509	8.01	26860	10.82	16157	14.65
07 PB149692BSD	8380	8.00	26080	10.82	15594	14.65

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.