



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005
Lab Code: CHEM Case No.: N6070 SAS No.: N6070 SDG No.: N6070
Instrument ID: BNA_N Calibration Date/Time: 12/19/2022 11:22
Lab File ID: BN023280.D Init. Calib. Date(s): 12/08/2022 12/08/2022
EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 14:00 17:40
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.678	0.625		-7.8	20.0
Fluoranthene-d10	0.936	0.950		1.5	20.0
2-Fluorophenol	0.745	0.926		24.3	20.0
Phenol-d6	0.947	1.181		24.7	20.0
Nitrobenzene-d5	0.264	0.271		2.7	20.0
2-Fluorobiphenyl	1.598	1.405		-12.1	20.0
2,4,6-Tribromophenol	0.145	0.172		18.6	20.0
Terphenyl-d14	0.649	0.685		5.5	20.0
1,4-Dioxane	0.395	0.440		11.4	20.0

All other compounds must meet a minimum RRF of 0.010.