



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: N6070 SAS No.: N6070 SDG NO.: N6070

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/19/2022

Lab File ID: BN023280.D Time Analyzed: 11:22

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	33722	17.39	25842	21.58	18763	24.027
UPPER LIMIT	67444	17.89	51684	22.08	37526	24.527
LOWER LIMIT	16861	16.89	12921	21.08	9381.5	23.527
EPA SAMPLE NO.						
01 PB149692BL	38498	17.39	26428	21.58	19388	24.03
02 GW-BR-04-226-245-121422-H	30735	17.39	24364	21.58	17711	24.03
03 PB149692BS	34289	17.39	26436	21.58	18732	24.03
04 PB149692BSD	33450	17.39	27783	21.58	19326	24.03

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.