



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120722

Review By	John Carlone	Review On	12/8/2022 10:16:51 AM				
Supervise By	Mahesh Dadoda	Supervise On	12/8/2022 10:31:45 AM				
SubDirectory	VX120722	HP Acquire Method	MSVOA_X	HP Processing Method	82X120722W.M		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		VP117343,VP117344 VP117345,VP117376,VP117347,VP117348,VP117349VP117350					
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP117352,VP117353 VP117351					

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX033272.D	07 Dec 2022 09:17		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX033273.D	07 Dec 2022 09:46		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX033274.D	07 Dec 2022 10:09		JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX033275.D	07 Dec 2022 10:32	com#13 fail	JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX033276.D	07 Dec 2022 10:55		JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX033277.D	07 Dec 2022 11:18		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX033278.D	07 Dec 2022 11:42		JC/MD	Ok,M
8	VIBLK	VIBLK	VX033279.D	07 Dec 2022 12:05		JC/MD	Ok
9	VSTDICV050	ICVVX120722	VX033280.D	07 Dec 2022 12:51	pH#Lot#V12668	JC/MD	Ok,M
10	BFB	BFB	VX033281.D	07 Dec 2022 14:57		JC/MD	Ok
11	VSTDCCC050	VSTDCCC050	VX033282.D	07 Dec 2022 15:25		JC/MD	Ok,M
12	VX1207MBL01	VX1207MBL01	VX033283.D	07 Dec 2022 15:53	Surrogate Fail	JC/MD	Not Ok
13	VX1207WBL01	VX1207WBL01	VX033284.D	07 Dec 2022 16:17		JC/MD	Ok
14	VX1207WBL02	VX1207WBL02	VX033285.D	07 Dec 2022 16:45		JC/MD	Ok
15	VX1207WBS01	VX1207WBS01	VX033286.D	07 Dec 2022 17:13		JC/MD	Ok,M
16	VX1207WBSD01	VX1207WBSD01	VX033287.D	07 Dec 2022 17:36		JC/MD	Ok,M
17	N5865-05	LEA-5SR	VX033288.D	07 Dec 2022 17:59	vial B pH<2	JC/MD	Ok
18	N5865-07	LEA-5D	VX033289.D	07 Dec 2022 18:22	vial B pH<2	JC/MD	Ok
19	N5875-03	MW-38	VX033290.D	07 Dec 2022 18:46	vial B pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120722

Review By	John Carlone	Review On	12/8/2022 10:16:51 AM					
Supervise By	Mahesh Dadoda	Supervise On	12/8/2022 10:31:45 AM					
SubDirectory	VX120722	HP Acquire Method	MSVOA_X	HP Processing Method	82X120722W.M			
STD. NAME		STD REF.#						
Tune/Reschk		VP117343,VP117344						
Initial Calibration Stds		VP117345,VP117376,VP117347,VP117348,VP117349VP117350						
CCC		VP117352,VP117353						
Internal Standard/PEM								
ICV/I.BLK		VP117351						
Surrogate Standard								
MS/MSD Standard								
LCS Standard								

20	N5875-06	MW-40	VX033291.D	07 Dec 2022 19:09	vial B pH<2	JC/MD	Ok
21	N5865-09DL	LEA-5IDL	VX033292.D	07 Dec 2022 19:32	vial B pH<2	JC/MD	Ok
22	N5865-11DL	DUP-01DL	VX033293.D	07 Dec 2022 19:55	vial B pH<2	JC/MD	Not Ok
23	N5865-01DL	LEA-7IDL	VX033294.D	07 Dec 2022 20:18	Confirm concentration for cis-1,2-Dichloroethene, not matched	JC/MD	Not Ok
24	VIBLK	VIBLK	VX033295.D	07 Dec 2022 20:41		JC/MD	Ok
25	N5912-01	FMC-TOTE-1207	VX033296.D	07 Dec 2022 21:05	vial A pH<2 Need 10x	JC/MD	Dilution
26	N5911-01	A3594TOG	VX033297.D	07 Dec 2022 21:28	vial A pH<2 E flag in previous sample	JC/MD	ReRun
27	N5906-01	MW-8	VX033298.D	07 Dec 2022 21:51	vial A pH<2	JC/MD	Ok
28	N5906-02	MW-12	VX033299.D	07 Dec 2022 22:14	vial A pH<2	JC/MD	Ok
29	N5906-03	MW-16	VX033300.D	07 Dec 2022 22:37	vial A pH<2	JC/MD	Ok
30	N5906-04	MW-20	VX033301.D	07 Dec 2022 23:00	vial A pH<2	JC/MD	Ok
31	N5906-05	H-3R	VX033302.D	07 Dec 2022 23:24	vial A pH<2	JC/MD	Ok
32	N5906-06	MW-3RR	VX033303.D	07 Dec 2022 23:47	vial A pH<2 Need 5X	JC/MD	Dilution
33	N5906-07	MW-13R	VX033304.D	08 Dec 2022 00:10	vial A pH<2 Need 10x	JC/MD	Dilution
34	N5906-08	MW-24	VX033305.D	08 Dec 2022 00:33	vial A pH<2	JC/MD	Ok
35	N5906-09	MW-28	VX033306.D	08 Dec 2022 00:56	vial A pH<2 Need 5X	JC/MD	Dilution
36	VSTDCCC050	VSTDCCC050EC	VX033307.D	08 Dec 2022 01:19		JC/MD	Ok,M

M : Manual Integration



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX121522

Review By	John Carlone	Review On	12/16/2022 9:45:30 AM				
Supervise By	Mahesh Dadoda	Supervise On	12/16/2022 12:26:32 PM				
SubDirectory	VX121522	HP Acquire Method	HP Processing Method		82X120722W.M		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		VP117542					
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP117544,VP117545					

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX033429.D	15 Dec 2022 09:39		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX033430.D	15 Dec 2022 10:10	pH#Lot#V12668	JC/MD	Ok,M
3	VX1215WBL01	VX1215WBL01	VX033431.D	15 Dec 2022 10:39		JC/MD	Ok
4	VX1215WBS01	VX1215WBS01	VX033432.D	15 Dec 2022 11:07		JC/MD	Ok,M
5	VX1215WBSD01	VX1215WBSD01	VX033433.D	15 Dec 2022 11:46		JC/MD	Ok,M
6	N6070-02	TB-01-121422	VX033434.D	15 Dec 2022 12:09	vial A pH<2 TB	JC/MD	Ok
7	N6070-01	GW-BR-04-226-245-12	VX033435.D	15 Dec 2022 12:32	vial A pH<2 Surrogate Fail	JC/MD	ReRun
8	N6070-03	OWBR-01-160-180-12	VX033436.D	15 Dec 2022 12:55	vial A pH<2	JC/MD	Ok
9	N6070-04	OWBR-02-160-180-12	VX033437.D	15 Dec 2022 13:18	vial A pH<2	JC/MD	Ok
10	N6070-05	OWBR-03-128-148-12	VX033438.D	15 Dec 2022 13:41	vial A pH<2	JC/MD	Ok
11	PB149669TB	PB149669TB	VX033439.D	15 Dec 2022 14:04	vial A pH#6.0	JC/MD	Ok
12	N6051-02	SP-591	VX033440.D	15 Dec 2022 14:27	vial A pH#6.0	JC/MD	Ok
13	N6070-01	GW-BR-04-226-245-12	VX033441.D	15 Dec 2022 14:50	vial B pH<2	JC/MD	Ok
14	PB149652TB	PB149652TB	VX033442.D	15 Dec 2022 15:13	vial A pH#5.0	JC/MD	Ok
15	PB149654TB	PB149654TB	VX033443.D	15 Dec 2022 15:36	vial A pH#5.0	JC/MD	Ok
16	N6065-02	MH-23	VX033444.D	15 Dec 2022 15:59	vial A pH#5.0	JC/MD	Ok
17	N6065-05	TP-S	VX033445.D	15 Dec 2022 16:22	vial A pH#5.0	JC/MD	Ok
18	N6067-02	WC-SOIL-20221214	VX033446.D	15 Dec 2022 16:45	vial A pH#5.0	JC/MD	Ok
19	N6067-04	WC-LIQUID-20221214	VX033447.D	15 Dec 2022 17:08	vial A pH#5.0	JC/MD	Ok



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX121522

Review By	John Carlone	Review On	12/16/2022 9:45:30 AM				
Supervise By	Maresh Dadoda	Supervise On	12/16/2022 12:26:32 PM				
SubDirectory	VX121522	HP Acquire Method	HP Processing Method		82X120722W.M		
STD. NAME	STD REF.#						
Tune/Reschk Initial Calibration Stds	VP117542						
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP117544,VP117545						
20	VSTDCCC050	VSTDCCC050EC	VX033448.D	15 Dec 2022 17:31		JC/MD	Ok,M

M : Manual Integration