



Surrogate Summary

SDG No.: N6070

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
N6070-01	GW-BR-04-226-245-121422	1,2-Dichloroethane-d4	50	48.8	98	70 (74)	130 (125)
		Dibromofluoromethane	50	49.3	99	70 (75)	130 (124)
		Toluene-d8	50	47.6	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.6	97	70 (64)	130 (133)
N6070-02	TB-01-121422	1,2-Dichloroethane-d4	50	46.7	93	70 (74)	130 (125)
		Dibromofluoromethane	50	47.0	94	70 (75)	130 (124)
		Toluene-d8	50	45.7	91	70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.2	88	70 (64)	130 (133)
N6070-03	OWBR-01-160-180-121422	1,2-Dichloroethane-d4	50	47.8	96	70 (74)	130 (125)
		Dibromofluoromethane	50	48.2	96	70 (75)	130 (124)
		Toluene-d8	50	46.7	93	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.6	91	70 (64)	130 (133)
N6070-04	OWBR-02-160-180-121422	1,2-Dichloroethane-d4	50	44.4	89	70 (74)	130 (125)
		Dibromofluoromethane	50	44.6	89	70 (75)	130 (124)
		Toluene-d8	50	43.7	87	70 (86)	130 (113)
		4-Bromofluorobenzene	50	42.0	84	70 (64)	130 (133)
N6070-05	OWBR-03-128-148-121422	1,2-Dichloroethane-d4	50	49.3	99	70 (74)	130 (125)
		Dibromofluoromethane	50	48.8	98	70 (75)	130 (124)
		Toluene-d8	50	47.3	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.6	93	70 (64)	130 (133)
VX1215WBL01	VX1215WBL01	1,2-Dichloroethane-d4	50	43.0	86	70 (74)	130 (125)
		Dibromofluoromethane	50	43.9	88	70 (75)	130 (124)
		Toluene-d8	50	42.9	86	70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.5	87	70 (64)	130 (133)
VX1215WBS01	VX1215WBS01	1,2-Dichloroethane-d4	50	48.2	96	70 (74)	130 (125)
		Dibromofluoromethane	50	51.4	103	70 (75)	130 (124)
		Toluene-d8	50	54.1	108	70 (86)	130 (113)
		4-Bromofluorobenzene	50	58.6	117	70 (64)	130 (133)
VX1215WBSD0	VX1215WBSD01	1,2-Dichloroethane-d4	50	43.0	86	70 (74)	130 (125)
		Dibromofluoromethane	50	46.0	92	70 (75)	130 (124)
		Toluene-d8	50	47.4	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.1	102	70 (64)	130 (133)

