



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: N6070

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Datafile : VX033432.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VX1215WBS01	Vinyl chloride	20	19.0	ug/L	95			70 (65)	130 (117)	
	1,1-Dichloroethene	20	14.9	ug/L	75			70 (74)	130 (110)	
	1,1-Dichloroethane	20	19.9	ug/L	100			70 (78)	130 (112)	
	cis-1,2-Dichloroethene	20	20.6	ug/L	103			70 (77)	130 (110)	
	1,1,1-Trichloroethane	20	21.7	ug/L	109			70 (80)	130 (108)	
	Benzene	20	21.3	ug/L	106			70 (82)	130 (109)	
	1,2-Dichloroethane	20	21.1	ug/L	106			70 (80)	130 (115)	
	Trichloroethene	20	22.2	ug/L	111			70 (77)	130 (113)	
	1,1,2-Trichloroethane	20	22.0	ug/L	110			70 (83)	130 (112)	
	Tetrachloroethene	20	22.0	ug/L	110			70 (67)	130 (123)	