



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: N6070 SAS No.: N6070 SDG NO.: N6070
Lab File ID: VX033430.D Date Analyzed: 12/15/2022
Instrument ID: MSVOA_X Time Analyzed: 10:10
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	139266	5.55	214327	6.76	205390	10.06
UPPER LIMIT	278532	6.05	428654	7.257	410780	10.555
LOWER LIMIT	69633	5.05	107164	6.257	102695	9.555
EPA SAMPLE NO.						
GW-BR-04-226-245-121422	115233	5.55	182643	6.76	164061	10.06
TB-01-121422	128174	5.56	204029	6.76	177272	10.06
OWBR-01-160-180-121422	112732	5.56	179781	6.76	158358	10.06
OWBR-02-160-180-121422	111282	5.55	176885	6.76	155365	10.06
OWBR-03-128-148-121422	117851	5.55	187379	6.76	166169	10.06
VX1215WBL01	143365	5.55	223781	6.76	198464	10.06
VX1215WBS01	144488	5.55	222709	6.76	207849	10.06
VX1215WBSD01	137295	5.55	210886	6.76	199474	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.