



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: N6070 SAS No.: N6070 SDG NO.: N6070
Lab File ID: VX033430.D Date Analyzed: 12/15/2022
Instrument ID: MSVOA_X Time Analyzed: 10:10
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	103647	12.024				
UPPER LIMIT	207294	12.524				
LOWER LIMIT	51823.5	11.524				
EPA SAMPLE NO.						
GW-BR-04-226-245-121422	77379	12.02				
TB-01-121422	78645	12.02				
OWBR-01-160-180-121422	73130	12.02				
OWBR-02-160-180-121422	69614	12.02				
OWBR-03-128-148-121422	76359	12.02				
VX1215WBL01	92099	12.02				
VX1215WBS01	106397	12.02				
VX1215WBSD01	101751	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.