

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033280.D
 Acq On : 07 Dec 2022 12:51
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX120722

Quant Time: Dec 08 01:41:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.349	0.347	0.6	108	0.00
3 P	Chloromethane	0.404	0.366	9.4	103	0.00
4 C	Vinyl Chloride	0.461	0.431	6.5#	103	0.00
5 T	Bromomethane	0.338	0.322	4.7	110	0.00
6 T	Chloroethane	0.350	0.330	5.7	100	0.00
7 T	Trichlorofluoromethane	0.866	0.840	3.0	105	0.00
8 T	Diethyl Ether	0.313	0.284	9.3	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.459	0.460	-0.2	106	0.00
10 T	Methyl Iodide	0.503	0.439	12.7	91	0.00
11 T	Tert butyl alcohol	0.098	0.083	15.3	106	0.00
12 CM	1,1-Dichloroethene	0.453	0.435	4.0#	105	0.00
13 T	Acrolein	0.106	0.097	8.5	98	0.00
14 T	Allyl chloride	0.724	0.695	4.0	104	0.00
15 T	Acrylonitrile	0.243	0.221	9.1	100	0.00
16 T	Acetone	0.237	0.206	13.1	102	0.00
17 T	Carbon Disulfide	1.163	1.114	4.2	104	0.00
18 T	Methyl Acetate	0.687	0.631	8.2	100	0.00
19 T	Methyl tert-butyl Ether	1.570	1.513	3.6	105	0.00
20 T	Methylene Chloride	0.529	0.483	8.7	105	0.00
21 T	trans-1,2-Dichloroethene	0.496	0.483	2.6	105	0.00
22 T	Diisopropyl ether	1.551	1.502	3.2	102	0.00
23 T	Vinyl Acetate	1.204	1.180	2.0	103	0.00
24 P	1,1-Dichloroethane	0.911	0.859	5.7	104	0.00
25 T	2-Butanone	0.341	0.317	7.0	100	0.00
26 T	2,2-Dichloropropane	0.670	0.667	0.4	105	0.00
27 T	cis-1,2-Dichloroethene	0.593	0.575	3.0	105	0.00
28 T	Bromochloromethane	0.381	0.366	3.9	100	0.00
29 T	Tetrahydrofuran	0.221	0.206	6.8	100	-0.01
30 C	Chloroform	1.012	0.961	5.0#	103	0.00
31 T	Cyclohexane	0.762	0.771	-1.2	106	0.00
32 T	1,1,1-Trichloroethane	0.870	0.867	0.3	104	0.00
33 S	1,2-Dichloroethane-d4	0.321	0.301	6.2	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.225	0.216	4.0	107	0.00
36 T	1,1-Dichloropropene	0.438	0.437	0.2	105	0.00
37 T	Ethyl Acetate	0.414	0.396	4.3	101	0.00
38 T	Carbon Tetrachloride	0.491	0.487	0.8	103	0.00
39 T	Methylcyclohexane	0.491	0.521	-6.1	107	0.00
40 TM	Benzene	1.278	1.251	2.1	104	0.00
41 T	Methacrylonitrile	0.228	0.223	2.2	100	0.00
42 TM	1,2-Dichloroethane	0.508	0.484	4.7	102	0.00
43 T	Isopropyl Acetate	0.667	0.649	2.7	102	0.00
44 TM	Trichloroethene	0.352	0.353	-0.3	105	0.00
45 C	1,2-Dichloropropane	0.327	0.318	2.8#	104	0.00
46 T	Dibromomethane	0.249	0.238	4.4	103	0.00
47 T	Bromodichloromethane	0.485	0.476	1.9	103	0.00
48 T	Methyl methacrylate	0.344	0.334	2.9	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033280.D
 Acq On : 07 Dec 2022 12:51
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX120722

Quant Time: Dec 08 01:41:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	103	0.00
50 S	Toluene-d8	0.470	0.471	-0.2	107	0.00
51 T	4-Methyl-2-Pentanone	0.422	0.414	1.9	101	0.00
52 CM	Toluene	0.827	0.831	-0.5#	103	0.00
53 T	t-1,3-Dichloropropene	0.473	0.489	-3.4	106	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.530	-2.5	105	0.00
55 T	1,1,2-Trichloroethane	0.347	0.340	2.0	103	0.00
56 T	Ethyl methacrylate	0.489	0.502	-2.7	104	0.00
57 T	1,3-Dichloropropane	0.558	0.553	0.9	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.224	0.226	-0.9	102	0.00
59 T	2-Hexanone	0.311	0.310	0.3	102	0.00
60 T	Dibromochloromethane	0.383	0.393	-2.6	105	0.00
61 T	1,2-Dibromoethane	0.370	0.374	-1.1	105	0.00
62 S	4-Bromofluorobenzene	0.360	0.379	-5.3	108	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.342	0.337	1.5	106	0.00
65 PM	Chlorobenzene	1.013	1.001	1.2	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.383	0.377	1.6	103	0.00
67 C	Ethyl Benzene	1.764	1.792	-1.6#	105	0.00
68 T	m/p-Xylenes	0.686	0.705	-2.8	106	0.00
69 T	o-Xylene	0.678	0.687	-1.3	104	0.00
70 T	Styrene	1.111	1.157	-4.1	105	0.00
71 P	Bromoform	0.303	0.307	-1.3	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	3.624	3.587	1.0	105	0.00
74 T	N-amyl acetate	1.236	1.251	-1.2	105	0.00
75 P	1,1,2,2-Tetrachloroethane	1.150	1.083	5.8	104	0.00
76 T	1,2,3-Trichloropropane	1.017	0.981	3.5	106	0.00
77 T	Bromobenzene	0.915	0.882	3.6	106	0.00
78 T	n-propylbenzene	4.047	4.115	-1.7	106	0.00
79 T	2-Chlorotoluene	2.501	2.431	2.8	105	0.00
80 T	1,3,5-Trimethylbenzene	3.017	3.030	-0.4	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.327	0.311	4.9	104	0.00
82 T	4-Chlorotoluene	2.908	2.827	2.8	106	0.00
83 T	tert-Butylbenzene	3.008	3.026	-0.6	106	0.00
84 T	1,2,4-Trimethylbenzene	3.011	2.999	0.4	106	0.00
85 T	sec-Butylbenzene	3.510	3.625	-3.3	107	0.00
86 T	p-Isopropyltoluene	3.026	3.082	-1.9	107	0.00
87 T	1,3-Dichlorobenzene	1.658	1.640	1.1	107	0.00
88 T	1,4-Dichlorobenzene	1.693	1.642	3.0	107	0.00
89 T	n-Butylbenzene	2.478	2.650	-6.9	110	0.00
90 T	Hexachloroethane	0.514	0.513	0.2	106	0.00
91 T	1,2-Dichlorobenzene	1.657	1.580	4.6	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.248	0.235	5.2	105	0.00
93 T	1,2,4-Trichlorobenzene	1.001	1.018	-1.7	111	0.00
94 T	Hexachlorobutadiene	0.401	0.415	-3.5	113	0.00
95 T	Naphthalene	3.339	3.382	-1.3	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033280.D
Acq On : 07 Dec 2022 12:51
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX120722

Quant Time: Dec 08 01:41:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.014	1.015	-0.1	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6