



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033272.D
 Acq On : 07 Dec 2022 09:17
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Title : SW846 8260
 Last Update : Wed Dec 07 12:43:54 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	7090	PASS
75	95	30	60	53.1	20095	PASS
95	95	100	100	100.0	37840	PASS
96	95	5	9	6.9	2623	PASS
173	174	0.00	2	0.9	270	PASS
174	95	50	100	77.0	29149	PASS
175	174	5	9	7.6	2227	PASS
176	174	95	101	97.6	28448	PASS
177	176	5	9	6.5	1841	PASS