



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: N6070

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Datafile : VX033433.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VX1215WBSD01	Vinyl chloride	20	17.6	ug/L	88	8		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	16.8	ug/L	84	11		70 (74)	130 (110)	20 (20)
	1,1-Dichloroethane	20	18.5	ug/L	93	7		70 (78)	130 (112)	20 (20)
	cis-1,2-Dichloroethene	20	18.7	ug/L	94	9		70 (77)	130 (110)	20 (20)
	1,1,1-Trichloroethane	20	19.8	ug/L	99	10		70 (80)	130 (108)	20 (20)
	Benzene	20	19.7	ug/L	99	7		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	19.8	ug/L	99	7		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	21.0	ug/L	105	6		70 (77)	130 (113)	20 (20)
	1,1,2-Trichloroethane	20	20.5	ug/L	103	7		70 (83)	130 (112)	20 (20)
	Tetrachloroethene	20	19.9	ug/L	100	10		70 (67)	130 (123)	20 (20)