



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Louis Berger U.S., Inc., A WSP Company

Project Name: NYCDDC Phase II SCI Arthur Kill Road CEQR

Project # N/A

Chemtech Project # O1232

Test Name: VOC-TCLVOA-10

A. Number of Samples and Date of Receipt:

10 Solid samples were received on 01/19/2023.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Diesel Range Organics, Gasoline Range Organics, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, Paint Filter, PCB, Pesticide-TCL, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-PAH, SVOC-TCL BNA -20, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS and VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_Y were done using GC column Rxi-624Sil MS, which is 30 meters, 0.25 mm id, 1.4 um df, Restek Cat. #13868. The Trap was supplied by Supelco, VOCARB 3000, ATOMAX XYZ Concentrator. The analysis of VOC-TCLVOA-10 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for B-P35B [1,2-Dichloroethane-d4 - 193%], B-P35BRE [1 and 2-Dichloroethane-d4 - 183%].

The Internal Standards Areas met the acceptable requirements except for B-P35B and B-P35BRE the failure sample in surrogate and Internal standard was reanalyzed to confirm the failure as per method and reported while,

For B-P19B VIAL A analyzed but internal standard fail as a corrective action VIAL B analyzed but did not purge therefore VIAL A reported as final analysis.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

The %RSD is greater than 15% in the Initial Calibration method (82Y011623S.M) for Methylene Chloride, 1,2-Dichloroethane-d4, Toluene-d8 these compounds are passing on Linear Regression.

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

This data Package has been revised to correct the Client ID of sample O1232-01, O1232-02, O1232-04, O1232-05, O1232-06, O1232-07, O1232-08, O1232-09 as per client request.

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____