



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Initial and Continuing Calibration Verification

Client: Louis Berger U.S., Inc., A WSP Company

SDG No.: O1232

Project: NYCDDC Phase II SCI Arthur Kill Road CEQR

RunNo.: LB123753

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.098	0.099	99	85-115	01/20/2023
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	01/20/2023
Sample ID: CCV2 Reactive Cyanide	mg/L	0.26	0.25	104	90-110	01/20/2023
Sample ID: CCV3 Reactive Cyanide	mg/L	0.26	0.25	104	90-110	01/20/2023



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Initial and Continuing Calibration Verification

Client: Louis Berger U.S., Inc., A WSP Company

SDG No.: O1232

Project: NYCDDC Phase II SCI Arthur Kill Road CEQR

RunNo.: LB123757

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Corrosivity		pH	6.99	7	100	90-110	01/20/2023
Sample ID:	CCV1						
Corrosivity		pH	2.02	2.00	101	90-110	01/20/2023
Sample ID:	CCV2						
Corrosivity		pH	12.03	12.00	100	90-110	01/20/2023